

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 409**

**CONTROL**



**(Exp.) 2019-01-05**  
(YYYY - MM - DD)

Rev 1

		<b>ABX Lyse</b>																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				MICROS		MICROS				MICROS							
		CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>	<b>2.1</b>		± 0.4	<b>7.5</b>	<b>7.5</b>	<b>7.6</b>		± 1.0	<b>19.9</b>	<b>19.9</b>	<b>20.3</b>		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.29</b>	<b>2.29</b>	<b>2.32</b>		± 0.15	<b>4.63</b>	<b>4.63</b>	<b>4.64</b>		± 0.18	<b>5.85</b>	<b>5.85</b>	<b>5.87</b>		± 0.25			
HB HGB	g/dl	<b>6.1</b>	<b>6.1</b>	<b>6.2</b>		± 0.4	<b>13.4</b>	<b>13.4</b>	<b>13.5</b>		± 0.5	<b>18.1</b>	<b>18.1</b>	<b>18.2</b>		± 0.7			
	g/l	<b>61</b>	<b>61</b>	<b>62</b>		± 4	<b>134</b>	<b>134</b>	<b>135</b>		± 5	<b>181</b>	<b>181</b>	<b>182</b>		± 7			
HT HCT	mmol/l	<b>3.79</b>	<b>3.79</b>	<b>3.85</b>		± 0.25	<b>8.32</b>	<b>8.32</b>	<b>8.38</b>		± 0.31	<b>11.24</b>	<b>11.24</b>	<b>11.30</b>		± 0.44			
	%	<b>16.3</b>	<b>16.3</b>	<b>16.7</b>		± 2.0	<b>37.5</b>	<b>37.5</b>	<b>37.6</b>		± 2.5	<b>51.5</b>	<b>51.5</b>	<b>51.7</b>		± 3.0			
VGM MCV	l/l	<b>0.163</b>	<b>0.163</b>	<b>0.167</b>		± 0.020	<b>0.375</b>	<b>0.375</b>	<b>0.376</b>		± 0.025	<b>0.515</b>	<b>0.515</b>	<b>0.517</b>		± 0.030			
	µm <sup>3</sup> ; fl	<b>71</b>	<b>71</b>	<b>72</b>		± 4	<b>81</b>	<b>81</b>	<b>81</b>		± 4	<b>88</b>	<b>88</b>	<b>88</b>		± 4			
TGMH MCH	pg	<b>26.6</b>	<b>26.6</b>	<b>26.7</b>		± 2.4	<b>28.9</b>	<b>28.9</b>	<b>29.1</b>		± 2.4	<b>30.9</b>	<b>30.9</b>	<b>31.0</b>		± 2.8			
	fmol	<b>1.65</b>	<b>1.65</b>	<b>1.66</b>		± 0.15	<b>1.80</b>	<b>1.80</b>	<b>1.81</b>		± 0.15	<b>1.92</b>	<b>1.92</b>	<b>1.93</b>		± 0.18			
CCMH MCHC	g/dl	<b>37.5</b>	<b>37.5</b>	<b>37.1</b>		± 3.0	<b>35.7</b>	<b>35.7</b>	<b>35.9</b>		± 3.0	<b>35.2</b>	<b>35.2</b>	<b>35.2</b>		± 3.0			
	g/l	<b>375</b>	<b>375</b>	<b>371</b>		± 30	<b>357</b>	<b>357</b>	<b>359</b>		± 30	<b>352</b>	<b>352</b>	<b>352</b>		± 30			
	mmol/l	<b>23.30</b>	<b>23.30</b>	<b>23.05</b>		± 1.86	<b>22.19</b>	<b>22.19</b>	<b>22.31</b>		± 1.86	<b>21.83</b>	<b>21.83</b>	<b>21.88</b>		± 1.86			
IDR RDW	%	<b>13.5</b>	<b>13.5</b>	<b>13.0</b>		± 3.0	<b>13.5</b>	<b>13.5</b>	<b>13.0</b>		± 3.0	<b>13.0</b>	<b>13.0</b>	<b>12.0</b>		± 3.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>76</b>	<b>76</b>	<b>80</b>		± 20	<b>255</b>	<b>255</b>	<b>257</b>		± 40	<b>500</b>	<b>500</b>	<b>505</b>		± 60			
VPM MPV	µm <sup>3</sup> ; fl	<b>8.3</b>	<b>8.3</b>	<b>7.5</b>		± 2.0	<b>8.1</b>	<b>8.1</b>	<b>7.3</b>		± 2.0	<b>7.9</b>	<b>7.9</b>	<b>7.2</b>		± 2.0			
LYMPHO	#	<b>1.4</b>	<b>1.4</b>	<b>1.3</b>		± 0.2	<b>2.6</b>	<b>2.6</b>	<b>2.5</b>		± 0.5	<b>3.2</b>	<b>3.2</b>	<b>3.2</b>		± 1.2			
	%	<b>65.0</b>	<b>65.0</b>	<b>63.0</b>		± 8.0	<b>34.0</b>	<b>34.0</b>	<b>33.5</b>		± 6.0	<b>16.0</b>	<b>16.0</b>	<b>16.0</b>		± 6.0			
MONO	#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>		± 0.2	<b>0.6</b>	<b>0.6</b>	<b>0.6</b>		± 0.4	<b>1.1</b>	<b>1.1</b>	<b>1.0</b>		± 1.0			
	%	<b>9.0</b>	<b>9.0</b>	<b>9.5</b>		± 7.0	<b>8.0</b>	<b>8.0</b>	<b>7.5</b>		± 5.0	<b>5.5</b>	<b>5.5</b>	<b>5.0</b>		± 5.0			
GRANULO	#	<b>0.5</b>	<b>0.5</b>	<b>0.6</b>		± 0.2	<b>4.4</b>	<b>4.4</b>	<b>4.5</b>		± 0.6	<b>15.6</b>	<b>15.6</b>	<b>16.0</b>		± 1.5			
	%	<b>26.0</b>	<b>26.0</b>	<b>27.5</b>		± 7.0	<b>58.0</b>	<b>58.0</b>	<b>59.0</b>		± 6.0	<b>78.5</b>	<b>78.5</b>	<b>79.0</b>		± 7.0			
CRP	mg/dl	<b>0.65</b>	<b>0.55</b>	N/A		± 0.30	<b>2.80</b>	<b>2.70</b>	N/A		± 0.70	<b>5.10</b>	<b>4.89</b>	N/A		± 1.00			
	mg/l	<b>6.5</b>	<b>5.5</b>	N/A		± 3.0	<b>28.0</b>	<b>27.0</b>	N/A		± 7.0	<b>51.0</b>	<b>48.9</b>	N/A		± 10.0			

Ref: TEMP-1070 Rev.37 FRONT/RECTO 9930089-A