

# ABX Minotrol CRP



**LOT**

**MC 401**

**CONTROL**



(Exp.) **2017-01-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse													
				CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE	
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P			
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>	<b>2.1</b>	± 0.4	<b>7.5</b>	<b>7.5</b>	<b>7.6</b>	± 1.0	<b>20.0</b>	<b>20.0</b>	<b>20.2</b>	± 2.2			
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.43</b>	<b>2.43</b>	<b>2.40</b>	± 0.15	<b>4.60</b>	<b>4.60</b>	<b>4.60</b>	± 0.18	<b>5.67</b>	<b>5.67</b>	<b>5.70</b>	± 0.25			
HB	HGB	g/dl	<b>6.3</b>	<b>6.3</b>	<b>6.2</b>	± 0.4	<b>13.6</b>	<b>13.6</b>	<b>13.5</b>	± 0.5	<b>18.0</b>	<b>18.0</b>	<b>17.9</b>	± 0.7			
		g/l	<b>63</b>	<b>63</b>	<b>62</b>	± 4	<b>136</b>	<b>136</b>	<b>135</b>	± 5	<b>180</b>	<b>180</b>	<b>179</b>	± 7			
HT	HCT	mmol/l	<b>3.91</b>	<b>3.91</b>	<b>3.85</b>	± 0.25	<b>8.45</b>	<b>8.45</b>	<b>8.38</b>	± 0.31	<b>11.18</b>	<b>11.18</b>	<b>11.12</b>	± 0.44			
		%	<b>17.0</b>	<b>17.0</b>	<b>16.8</b>	± 2.0	<b>37.7</b>	<b>37.7</b>	<b>37.3</b>	± 2.5	<b>49.9</b>	<b>49.9</b>	<b>50.2</b>	± 3.0			
VGM	MCV	l/l	<b>0.170</b>	<b>0.170</b>	<b>0.168</b>	± 0.020	<b>0.377</b>	<b>0.377</b>	<b>0.373</b>	± 0.025	<b>0.499</b>	<b>0.499</b>	<b>0.502</b>	± 0.030			
		µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>	<b>70</b>	± 4	<b>82</b>	<b>82</b>	<b>81</b>	± 4	<b>88</b>	<b>88</b>	<b>88</b>	± 4			
TGMH	MCH	pg	<b>25.9</b>	<b>25.9</b>	<b>25.8</b>	± 2.4	<b>29.6</b>	<b>29.6</b>	<b>29.3</b>	± 2.4	<b>31.7</b>	<b>31.7</b>	<b>31.4</b>	± 2.8			
		fmol	<b>1.61</b>	<b>1.61</b>	<b>1.60</b>	± 0.15	<b>1.84</b>	<b>1.84</b>	<b>1.82</b>	± 0.15	<b>1.97</b>	<b>1.97</b>	<b>1.95</b>	± 0.18			
CCMH	MCHC	g/dl	<b>37.0</b>	<b>37.0</b>	<b>36.9</b>	± 3.0	<b>36.1</b>	<b>36.1</b>	<b>36.2</b>	± 3.0	<b>36.1</b>	<b>36.1</b>	<b>35.7</b>	± 3.0			
		g/l	<b>370</b>	<b>370</b>	<b>369</b>	± 30	<b>361</b>	<b>361</b>	<b>362</b>	± 30	<b>361</b>	<b>361</b>	<b>357</b>	± 30			
		mmol/l	<b>23.00</b>	<b>23.00</b>	<b>22.92</b>	± 1.86	<b>22.39</b>	<b>22.39</b>	<b>22.50</b>	± 1.86	<b>22.40</b>	<b>22.40</b>	<b>22.16</b>	± 1.86			
IDR	RDW	%	<b>14.0</b>	<b>14.0</b>	<b>13.5</b>	± 3.0	<b>14.0</b>	<b>14.0</b>	<b>14.0</b>	± 3.0	<b>13.0</b>	<b>13.0</b>	<b>13.0</b>	± 3.0			
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>75</b>	<b>75</b>	<b>74</b>	± 20	<b>246</b>	<b>246</b>	<b>247</b>	± 40	<b>480</b>	<b>480</b>	<b>484</b>	± 60			
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.0</b>	<b>8.0</b>	<b>7.0</b>	± 2.0	<b>7.8</b>	<b>7.8</b>	<b>7.0</b>	± 2.0	<b>7.5</b>	<b>7.5</b>	<b>7.0</b>	± 2.0			
LYMPHO	#	#	<b>1.3</b>	<b>1.3</b>	<b>1.3</b>	± 0.3	<b>2.4</b>	<b>2.4</b>	<b>2.4</b>	± 0.5	<b>3.1</b>	<b>3.1</b>	<b>3.1</b>	± 1.3			
		%	<b>63.0</b>	<b>63.0</b>	<b>62.0</b>	± 8.0	<b>32.5</b>	<b>32.5</b>	<b>32.0</b>	± 6.0	<b>15.5</b>	<b>15.5</b>	<b>15.5</b>	± 6.0			
MONO	#	#	<b>0.3</b>	<b>0.3</b>	<b>0.3</b>	± 0.2	<b>0.6</b>	<b>0.6</b>	<b>0.5</b>	± 0.4	<b>1.1</b>	<b>1.1</b>	<b>1.1</b>	± 1.0			
		%	<b>12.0</b>	<b>12.0</b>	<b>12.0</b>	± 7.0	<b>7.5</b>	<b>7.5</b>	<b>7.0</b>	± 5.0	<b>5.5</b>	<b>5.5</b>	<b>5.5</b>	± 5.0			
GRANULO	#	#	<b>0.5</b>	<b>0.5</b>	<b>0.5</b>	± 0.3	<b>4.5</b>	<b>4.5</b>	<b>4.6</b>	± 0.5	<b>15.8</b>	<b>15.8</b>	<b>16.0</b>	± 1.5			
		%	<b>25.0</b>	<b>25.0</b>	<b>26.0</b>	± 8.0	<b>60.0</b>	<b>60.0</b>	<b>61.0</b>	± 6.0	<b>79.0</b>	<b>79.0</b>	<b>79.0</b>	± 7.0			
CRP	mg/dl	mg/dl	<b>0.65</b>	<b>0.48</b>	N/A	± 0.30	<b>2.40</b>	<b>2.36</b>	N/A	± 0.70	<b>4.09</b>	<b>4.20</b>	N/A	± 1.00			
		mg/l	<b>6.5</b>	<b>4.8</b>	N/A	± 3.0	<b>24.0</b>	<b>23.6</b>	N/A	± 7.0	<b>40.9</b>	<b>42.0</b>	N/A	± 10.0			

Ref: TEMP-1070 Rev.36 FRONT/RECTO 9930089-A