

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT

MC 405

CONTROL



(Exp.) 2018-01-05
(YYYY - MM - DD)

Rev 1

| | | ABX Lyse | | | | | | | | | | | | | | | | | |
|--------------------------|---|-----------------|---------|-----------------|--|---------|-------------------------|---------|---------|-----------------|---------|-------|-------------------------|---------|---------|-----------------|--|---|-------------------------|
| PARAMETRES PARAMETERS | UNITES UNITS | CONTROL | | | | 1 | TOLERANCES TOLERANCE | CONTROL | | | | 2 | TOLERANCES TOLERANCE | CONTROL | | | | 3 | TOLERANCES TOLERANCE |
| | | MICROS | | | | | | MICROS | | | | | | MICROS | | | | | |
| | | CRP | CRP 200 | 60 8P 60 18P | | | | CRP | CRP 200 | 60 8P 60 18P | | | | CRP | CRP 200 | 60 8P 60 18P | | | |
| | | | | | | | | | | | | | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.1 | 2.1 | 2.2 | | ± 0.4 | 7.6 | 7.6 | 7.6 | | ± 1.0 | 20.2 | 20.2 | 20.3 | | ± 2.2 | | | |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.21 | 2.21 | 2.23 | | ± 0.15 | 4.67 | 4.67 | 4.67 | | ± 0.18 | 5.64 | 5.64 | 5.69 | | ± 0.25 | | | |
| HB HGB | g/dl | 5.8 | 5.8 | 5.8 | | ± 0.4 | 13.5 | 13.5 | 13.6 | | ± 0.5 | 17.7 | 17.7 | 17.8 | | ± 0.7 | | | |
| | g/l | 58 | 58 | 58 | | ± 4 | 135 | 135 | 136 | | ± 5 | 177 | 177 | 178 | | ± 7 | | | |
| HT HCT | mmol/l | 3.60 | 3.60 | 3.60 | | ± 0.25 | 8.38 | 8.38 | 8.45 | | ± 0.31 | 10.99 | 10.99 | 11.05 | | ± 0.44 | | | |
| | % | 15.5 | 15.5 | 15.6 | | ± 2.0 | 37.8 | 37.8 | 37.8 | | ± 2.5 | 49.6 | 49.6 | 50.6 | | ± 3.0 | | | |
| VGM MCV | l/l | 0.155 | 0.155 | 0.156 | | ± 0.020 | 0.378 | 0.378 | 0.378 | | ± 0.025 | 0.496 | 0.496 | 0.506 | | ± 0.030 | | | |
| | µm ³ ; fl | 70 | 70 | 70 | | ± 4 | 81 | 81 | 81 | | ± 4 | 88 | 88 | 89 | | ± 4 | | | |
| TGMH MCH | pg | 26.2 | 26.2 | 26.0 | | ± 2.4 | 28.9 | 28.9 | 29.1 | | ± 2.4 | 31.4 | 31.4 | 31.3 | | ± 2.8 | | | |
| | fmol | 1.63 | 1.63 | 1.62 | | ± 0.15 | 1.80 | 1.80 | 1.81 | | ± 0.15 | 1.95 | 1.95 | 1.94 | | ± 0.18 | | | |
| CCMH MCHC | g/dl | 37.5 | 37.5 | 37.2 | | ± 3.0 | 35.7 | 35.7 | 36.0 | | ± 3.0 | 35.7 | 35.7 | 35.1 | | ± 3.0 | | | |
| | g/l | 375 | 375 | 372 | | ± 30 | 357 | 357 | 360 | | ± 30 | 357 | 357 | 351 | | ± 30 | | | |
| | mmol/l | 23.28 | 23.28 | 23.07 | | ± 1.86 | 22.16 | 22.16 | 22.33 | | ± 1.86 | 22.15 | 22.15 | 21.83 | | ± 1.86 | | | |
| IDR RDW | % | 14.5 | 14.5 | 13.0 | | ± 3.0 | 14.0 | 14.0 | 13.5 | | ± 3.0 | 13.5 | 13.5 | 12.5 | | ± 3.0 | | | |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 69 | 69 | 71 | | ± 20 | 255 | 255 | 257 | | ± 40 | 530 | 530 | 535 | | ± 60 | | | |
| VPM MPV | µm ³ ; fl | 8.7 | 8.7 | 8.0 | | ± 2.0 | 8.1 | 8.1 | 7.7 | | ± 2.0 | 8.3 | 8.3 | 7.8 | | ± 2.0 | | | |
| LYMPHO | # | 1.4 | 1.4 | 1.5 | | ± 0.2 | 2.7 | 2.7 | 2.7 | | ± 0.5 | 3.2 | 3.2 | 3.2 | | ± 1.3 | | | |
| | % | 67.5 | 67.5 | 66.0 | | ± 8.0 | 35.0 | 35.0 | 35.0 | | ± 6.0 | 16.0 | 16.0 | 16.0 | | ± 6.0 | | | |
| MONO | # | 0.2 | 0.2 | 0.2 | | ± 0.2 | 0.5 | 0.5 | 0.5 | | ± 0.4 | 1.1 | 1.1 | 1.0 | | ± 1.0 | | | |
| | % | 8.5 | 8.5 | 8.5 | | ± 7.0 | 7.0 | 7.0 | 6.5 | | ± 5.0 | 5.5 | 5.5 | 5.0 | | ± 5.0 | | | |
| GRANULO | # | 0.6 | 0.6 | 0.6 | | ± 0.2 | 4.4 | 4.4 | 4.4 | | ± 0.5 | 15.9 | 15.9 | 16.0 | | ± 1.5 | | | |
| | % | 24.0 | 24.0 | 25.5 | | ± 7.0 | 58.0 | 58.0 | 58.5 | | ± 6.0 | 78.5 | 78.5 | 79.0 | | ± 7.0 | | | |
| CRP | mg/dl | 0.61 | 0.46 | N/A | | ± 0.30 | 2.24 | 1.94 | N/A | | ± 0.70 | 3.69 | 3.27 | N/A | | ± 1.00 | | | |
| | mg/l | 6.1 | 4.6 | N/A | | ± 3.0 | 22.4 | 19.4 | N/A | | ± 7.0 | 36.9 | 32.7 | N/A | | ± 10.0 | | | |

Ref: TEMP-1070 Rev.37 FRONT/RECTO 9930089-A