

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 074**

**CONTROL**



(Exp.) **2014-10-05**

(YYYY - MM - DD)

Rev 1

		<b>ABX Lyse</b>																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					1	TOLERANCES TOLERANCE	CONTROL					2	TOLERANCES TOLERANCE	CONTROL					3	TOLERANCES TOLERANCE
		MICROS							MICROS							MICROS						
		CRP	CRP 200	60 8P 60 18P					CRP	CRP 200	60 8P 60 18P					CRP	CRP 200	60 8P 60 18P				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>	<b>2.1</b>			± 0.4	<b>7.4</b>	<b>7.4</b>	<b>7.6</b>			± 1.0	<b>19.7</b>	<b>19.7</b>	<b>20.1</b>			± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.34</b>	<b>2.34</b>	<b>2.36</b>			± 0.15	<b>4.59</b>	<b>4.59</b>	<b>4.64</b>			± 0.18	<b>5.47</b>	<b>5.47</b>	<b>5.53</b>			± 0.25			
HB HGB	g/dl	<b>6.0</b>	<b>6.0</b>	<b>6.0</b>			± 0.4	<b>13.4</b>	<b>13.4</b>	<b>13.4</b>			± 0.5	<b>17.7</b>	<b>17.7</b>	<b>17.7</b>			± 0.7			
	g/l	<b>60</b>	<b>60</b>	<b>60</b>			± 4	<b>134</b>	<b>134</b>	<b>134</b>			± 5	<b>177</b>	<b>177</b>	<b>177</b>			± 7			
	mmol/l	<b>3.73</b>	<b>3.73</b>	<b>3.73</b>			± 0.25	<b>8.32</b>	<b>8.32</b>	<b>8.32</b>			± 0.31	<b>10.99</b>	<b>10.99</b>	<b>10.99</b>			± 0.44			
HT HCT	%	<b>16.4</b>	<b>16.4</b>	<b>16.3</b>			± 2.0	<b>37.2</b>	<b>37.2</b>	<b>37.6</b>			± 2.5	<b>49.2</b>	<b>49.2</b>	<b>49.8</b>			± 3.0			
	l/l	<b>0.164</b>	<b>0.164</b>	<b>0.163</b>			± 0.020	<b>0.372</b>	<b>0.372</b>	<b>0.376</b>			± 0.025	<b>0.492</b>	<b>0.492</b>	<b>0.498</b>			± 0.030			
VGM MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>	<b>69</b>			± 4	<b>81</b>	<b>81</b>	<b>81</b>			± 4	<b>90</b>	<b>90</b>	<b>90</b>			± 4			
TGMH MCH	pg	<b>25.6</b>	<b>25.6</b>	<b>25.4</b>			± 2.4	<b>29.2</b>	<b>29.2</b>	<b>28.9</b>			± 2.4	<b>32.4</b>	<b>32.4</b>	<b>32.0</b>			± 2.8			
	fmol	<b>1.59</b>	<b>1.59</b>	<b>1.58</b>			± 0.15	<b>1.81</b>	<b>1.81</b>	<b>1.79</b>			± 0.15	<b>2.01</b>	<b>2.01</b>	<b>1.99</b>			± 0.18			
CCMH MCHC	g/dl	<b>36.6</b>	<b>36.6</b>	<b>36.8</b>			± 3.0	<b>36.0</b>	<b>36.0</b>	<b>35.7</b>			± 3.0	<b>36.0</b>	<b>36.0</b>	<b>35.6</b>			± 3.0			
	g/l	<b>366</b>	<b>366</b>	<b>368</b>			± 30	<b>360</b>	<b>360</b>	<b>357</b>			± 30	<b>360</b>	<b>360</b>	<b>356</b>			± 30			
	mmol/l	<b>22.75</b>	<b>22.75</b>	<b>22.88</b>			± 1.86	<b>22.38</b>	<b>22.38</b>	<b>22.14</b>			± 1.86	<b>22.33</b>	<b>22.33</b>	<b>22.08</b>			± 1.86			
IDR RDW	%	<b>14.0</b>	<b>14.0</b>	<b>14.0</b>			± 3.0	<b>13.7</b>	<b>13.7</b>	<b>13.7</b>			± 3.0	<b>13.0</b>	<b>13.0</b>	<b>13.2</b>			± 3.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>80</b>	<b>80</b>	<b>80</b>			± 20	<b>257</b>	<b>257</b>	<b>258</b>			± 40	<b>495</b>	<b>495</b>	<b>500</b>			± 60			
VPM MPV	µm <sup>3</sup> ; fl	<b>8.6</b>	<b>8.6</b>	<b>8.0</b>			± 2.0	<b>7.7</b>	<b>7.7</b>	<b>7.4</b>			± 2.0	<b>7.6</b>	<b>7.6</b>	<b>7.3</b>			± 2.0			
LYMPHO	#	<b>1.3</b>	<b>1.3</b>	<b>1.3</b>			± 0.2	<b>2.5</b>	<b>2.5</b>	<b>2.5</b>			± 0.5	<b>3.1</b>	<b>3.1</b>	<b>3.1</b>			± 1.1			
	%	<b>63.5</b>	<b>63.5</b>	<b>63.0</b>			± 8.0	<b>33.5</b>	<b>33.5</b>	<b>33.5</b>			± 6.0	<b>15.5</b>	<b>15.5</b>	<b>15.5</b>			± 5.0			
MONO	#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>			± 0.2	<b>0.6</b>	<b>0.6</b>	<b>0.6</b>			± 0.4	<b>1.0</b>	<b>1.0</b>	<b>1.0</b>			± 1.0			
	%	<b>8.5</b>	<b>8.5</b>	<b>8.5</b>			± 7.0	<b>8.0</b>	<b>8.0</b>	<b>7.5</b>			± 5.0	<b>5.0</b>	<b>5.0</b>	<b>5.0</b>			± 5.0			
GRANULO	#	<b>0.6</b>	<b>0.6</b>	<b>0.6</b>			± 0.2	<b>4.3</b>	<b>4.3</b>	<b>4.5</b>			± 0.5	<b>15.7</b>	<b>15.7</b>	<b>16.0</b>			± 1.5			
	%	<b>28.0</b>	<b>28.0</b>	<b>28.5</b>			± 7.0	<b>58.5</b>	<b>58.5</b>	<b>59.0</b>			± 6.0	<b>79.5</b>	<b>79.5</b>	<b>79.5</b>			± 7.0			
CRP	mg/dl	<b>0.81</b>	<b>0.46</b>	N/A			± 0.30	<b>2.84</b>	<b>2.32</b>	N/A			± 0.70	<b>4.46</b>	<b>3.85</b>	N/A			± 1.00			
	mg/l	<b>8.1</b>	<b>4.6</b>	N/A			± 3.0	<b>28.4</b>	<b>23.2</b>	N/A			± 7.0	<b>44.6</b>	<b>38.5</b>	N/A			± 10.0			

9930089-A  
FRONT/RECTO  
Ref: TEMP-1070 Rev.35