

LOT MC 100
 Rev 2

CONTROL

(Exp.) 2011-01-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0	2.0	± 0.4	7.5	7.5	7.5	± 1.0	20.0	20.0	20.1	± 2.2			
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.30	2.30	2.34	± 0.15	4.43	4.43	4.48	± 0.18	5.50	5.50	5.58	± 0.25			
HB	HGB	g/dl	6.0	6.0	5.9	± 0.4	13.4	13.4	13.3	± 0.5	17.4	17.4	17.5	± 0.7			
		g/l	60	60	59	± 4	134	134	133	± 5	174	174	175	± 7			
		mmol/l	3.73	3.73	3.66	± 0.25	8.32	8.32	8.26	± 0.31	10.81	10.81	10.87	± 0.44			
HT	HCT	%	16.3	16.3	16.6	± 2.0	37.2	37.2	37.2	± 2.5	49.0	49.0	49.7	± 3.0			
		l/l	0.163	0.163	0.166	± 0.020	0.372	0.372	0.372	± 0.025	0.490	0.490	0.497	± 0.030			
VGM	MCV	µm ³ ; fl	71	71	71	± 4	84	84	83	± 4	89	89	89	± 4			
TGMH	MCH	pg	26.1	26.1	25.2	± 2.4	30.2	30.2	29.7	± 2.4	31.6	31.6	31.4	± 2.8			
		fmol	1.62	1.62	1.57	± 0.15	1.88	1.88	1.84	± 0.15	1.96	1.96	1.95	± 0.18			
CCMH	MCHC	g/dl	36.7	36.7	35.5	± 3.0	36.0	36.0	35.8	± 3.0	35.5	35.5	35.2	± 3.0			
		g/l	367	367	355	± 30	360	360	358	± 30	355	355	352	± 30			
		mmol/l	22.82	22.82	22.05	± 1.86	22.36	22.36	22.21	± 1.86	22.07	22.07	21.88	± 1.86			
IDR	RDW	%	13.6	13.6	13.5	± 3.0	12.8	12.8	13.0	± 3.0	12.2	12.2	12.4	± 3.0			
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	80	80	82	± 20	245	245	250	± 40	495	495	500	± 60			
VPM	MPV	µm ³ ; fl	8.5	8.5	8.5	± 2.0	7.9	7.9	7.9	± 2.0	7.9	7.9	8.0	± 2.0			
LYMPHO		#	1.2	1.2	1.3	± 0.2	2.5	2.5	2.6	± 0.5	3.0	3.0	3.0	± 1.0			
		%	62.0	62.0	63.5	± 7.0	33.5	33.5	34.0	± 6.0	15.0	15.0	15.0	± 5.0			
MONO		#	0.2	0.2	0.2	± 0.2	0.5	0.5	0.5	± 0.5	1.0	1.0	1.0	± 1.0			
		%	11.0	11.0	10.0	± 7.0	7.2	7.2	7.0	± 5.0	5.0	5.0	5.0	± 5.0			
GRANULO		#	0.5	0.5	0.5	± 0.2	4.4	4.4	4.4	± 0.5	16.0	16.0	16.1	± 1.5			
		%	27.0	27.0	26.5	± 7.0	59.3	59.3	59.0	± 6.0	80.0	80.0	80.0	± 7.0			
CRP		mg/dl	0.58	0.55	N/A	± 0.30	2.20	2.30	N/A	± 0.70	4.20	4.60	N/A	± 1.00			
		mg/l	5.8	5.5	N/A	± 3.0	22.0	23.0	N/A	± 7.0	42.0	46.0	N/A	± 10.0			

Ref. FORM-1070 Rev.12 FRONT / RECTO 9930045-B