

# ABX Minotrol CRP



**LOT**

**MC 044**

**CONTROL**



**(Exp.) 2014-07-05**

Rev 2

(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse													TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3			
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE			
				CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P				
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>	<b>2.0</b>		± 0.4	<b>7.4</b>	<b>7.4</b>	<b>7.5</b>		± 1.0	<b>20.1</b>	<b>20.1</b>	<b>20.5</b>		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.37</b>	<b>2.37</b>	<b>2.39</b>		± 0.15	<b>4.60</b>	<b>4.60</b>	<b>4.64</b>		± 0.18	<b>5.66</b>	<b>5.66</b>	<b>5.72</b>		± 0.25	
HB	HGB	g/dl	<b>6.1</b>	<b>6.1</b>	<b>6.0</b>		± 0.4	<b>13.5</b>	<b>13.5</b>	<b>13.5</b>		± 0.5	<b>17.9</b>	<b>17.9</b>	<b>17.9</b>		± 0.7	
		g/l	<b>61</b>	<b>61</b>	<b>60</b>		± 4	<b>135</b>	<b>135</b>	<b>135</b>		± 5	<b>179</b>	<b>179</b>	<b>179</b>		± 7	
		mmol/l	<b>3.79</b>	<b>3.79</b>	<b>3.73</b>		± 0.25	<b>8.38</b>	<b>8.38</b>	<b>8.38</b>		± 0.31	<b>11.12</b>	<b>11.12</b>	<b>11.12</b>		± 0.44	
HT	HCT	%	<b>16.4</b>	<b>16.4</b>	<b>16.5</b>		± 2.0	<b>37.3</b>	<b>37.3</b>	<b>37.6</b>		± 2.5	<b>50.4</b>	<b>50.4</b>	<b>50.9</b>		± 3.0	
		l/l	<b>0.164</b>	<b>0.164</b>	<b>0.165</b>		± 0.020	<b>0.373</b>	<b>0.373</b>	<b>0.376</b>		± 0.025	<b>0.504</b>	<b>0.504</b>	<b>0.509</b>		± 0.030	
VGM	MCV	µm <sup>3</sup> ; fl	<b>69</b>	<b>69</b>	<b>69</b>		± 4	<b>81</b>	<b>81</b>	<b>81</b>		± 4	<b>89</b>	<b>89</b>	<b>89</b>		± 4	
TGMH	MCH	pg	<b>25.7</b>	<b>25.7</b>	<b>25.1</b>		± 2.4	<b>29.3</b>	<b>29.3</b>	<b>29.1</b>		± 2.4	<b>31.6</b>	<b>31.6</b>	<b>31.3</b>		± 2.8	
		fmol	<b>1.60</b>	<b>1.60</b>	<b>1.56</b>		± 0.15	<b>1.82</b>	<b>1.82</b>	<b>1.81</b>		± 0.15	<b>1.96</b>	<b>1.96</b>	<b>1.94</b>		± 0.18	
CCMH	MCHC	g/dl	<b>37.3</b>	<b>37.3</b>	<b>36.4</b>		± 3.0	<b>36.2</b>	<b>36.2</b>	<b>35.9</b>		± 3.0	<b>35.5</b>	<b>35.5</b>	<b>35.2</b>		± 3.0	
		g/l	<b>373</b>	<b>373</b>	<b>364</b>		± 30	<b>362</b>	<b>362</b>	<b>359</b>		± 30	<b>355</b>	<b>355</b>	<b>352</b>		± 30	
		mmol/l	<b>23.16</b>	<b>23.16</b>	<b>22.59</b>		± 1.86	<b>22.50</b>	<b>22.50</b>	<b>22.31</b>		± 1.86	<b>22.07</b>	<b>22.07</b>	<b>21.84</b>		± 1.86	
IDR	RDW	%	<b>13.5</b>	<b>13.5</b>	<b>13.5</b>		± 3.0	<b>13.4</b>	<b>13.4</b>	<b>13.6</b>		± 3.0	<b>13.2</b>	<b>13.2</b>	<b>13.5</b>		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>79</b>	<b>79</b>	<b>79</b>		± 20	<b>255</b>	<b>255</b>	<b>255</b>		± 40	<b>500</b>	<b>500</b>	<b>505</b>		± 60	
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.8</b>	<b>8.8</b>	<b>8.4</b>		± 2.0	<b>8.1</b>	<b>8.1</b>	<b>7.7</b>		± 2.0	<b>7.8</b>	<b>7.8</b>	<b>7.5</b>		± 2.0	
LYMPHO		#	<b>1.3</b>	<b>1.3</b>	<b>1.3</b>		± 0.3	<b>2.5</b>	<b>2.5</b>	<b>2.6</b>		± 0.5	<b>3.0</b>	<b>3.0</b>	<b>3.1</b>		± 1.3	
		%	<b>65.5</b>	<b>65.5</b>	<b>64.5</b>		± 8.0	<b>34.0</b>	<b>34.0</b>	<b>34.0</b>		± 6.0	<b>15.0</b>	<b>15.0</b>	<b>15.0</b>		± 6.0	
MONO		#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>		± 0.2	<b>0.6</b>	<b>0.6</b>	<b>0.6</b>		± 0.4	<b>1.4</b>	<b>1.4</b>	<b>1.1</b>		± 1.0	
		%	<b>9.5</b>	<b>9.5</b>	<b>9.0</b>		± 7.0	<b>8.5</b>	<b>8.5</b>	<b>7.5</b>		± 5.0	<b>7.0</b>	<b>7.0</b>	<b>5.5</b>		± 5.0	
GRANULO		#	<b>0.5</b>	<b>0.5</b>	<b>0.5</b>		± 0.2	<b>4.3</b>	<b>4.3</b>	<b>4.4</b>		± 0.5	<b>15.7</b>	<b>15.7</b>	<b>16.3</b>		± 1.5	
		%	<b>25.0</b>	<b>25.0</b>	<b>26.5</b>		± 7.0	<b>57.5</b>	<b>57.5</b>	<b>58.5</b>		± 6.0	<b>78.0</b>	<b>78.0</b>	<b>79.5</b>		± 7.0	
CRP		mg/dl	<b>0.90</b>	<b>0.50</b>	N/A		± 0.30	<b>2.90</b>	<b>2.50</b>	N/A		± 0.70	<b>4.60</b>	<b>4.10</b>	N/A		± 1.00	
		mg/l	<b>9.0</b>	<b>5.0</b>	N/A		± 3.0	<b>29.0</b>	<b>25.0</b>	N/A		± 7.0	<b>46.0</b>	<b>41.0</b>	N/A		± 10.0	

9930074-A FRONT/RECTO Ref: TEMP-1070 Rev.34