

# ABX Minotrol CRP



**LOT** MC 072  
Rev 1

**CONTROL**

(Exp.) 2012-10-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse																
				CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE				
				MICROS			MICROS			MICROS										
				CRP	CRP 200	60 8P 60 18P	TOLERANCES TOLERANCE	CRP	CRP 200	60 8P 60 18P	TOLERANCES TOLERANCE	CRP	CRP 200	60 8P 60 18P	TOLERANCES TOLERANCE					
GB	WBC	$10^9/mm^3 ; 10^9/l$	1.9	1.9	2.0			$\pm 0.4$	7.4	7.4	7.5			$\pm 1.0$	19.7	19.7	19.9			$\pm 2.2$
GR	RBC	$10^9/mm^3 ; 10^{12}/l$	2.40	2.40	2.38			$\pm 0.15$	4.70	4.70	4.68			$\pm 0.18$	5.59	5.59	5.65			$\pm 0.25$
HB	HGB	g/dl	6.2	6.2	6.2			$\pm 0.4$	13.6	13.6	13.6			$\pm 0.5$	17.8	17.8	17.9			$\pm 0.7$
		g/l	62	62	62			$\pm 4$	136	136	136			$\pm 5$	178	178	179			$\pm 7$
		mmol/l	3.85	3.85	3.85			$\pm 0.25$	8.45	8.45	8.45			$\pm 0.31$	11.05	11.05	11.12			$\pm 0.44$
HT	HCT	%	17.3	17.3	17.1			$\pm 2.0$	38.5	38.5	38.4			$\pm 2.5$	50.3	50.3	50.9			$\pm 3.0$
		l/l	0.173	0.173	0.171			$\pm 0.020$	0.385	0.385	0.384			$\pm 0.025$	0.503	0.503	0.509			$\pm 0.030$
VGM	MCV	$\mu m^3 ; fl$	72	72	72			$\pm 4$	82	82	82			$\pm 4$	90	90	90			$\pm 4$
TGMH	MCH	pg	25.8	25.8	26.1			$\pm 2.4$	28.9	28.9	29.1			$\pm 2.4$	31.8	31.8	31.7			$\pm 2.8$
		fmol	1.60	1.60	1.62			$\pm 0.15$	1.80	1.80	1.80			$\pm 0.15$	1.98	1.98	1.97			$\pm 0.18$
CCMH	MCHC	g/dl	35.9	35.9	36.2			$\pm 3.0$	35.3	35.3	35.4			$\pm 3.0$	35.4	35.4	35.2			$\pm 3.0$
		g/l	359	359	362			$\pm 30$	353	353	354			$\pm 30$	354	354	352			$\pm 30$
		mmol/l	22.28	22.28	22.47			$\pm 1.86$	21.91	21.91	22.01			$\pm 1.86$	21.97	21.97	21.86			$\pm 1.86$
IDR	RDW	%	14.0	14.0	13.5			$\pm 3.0$	13.2	13.2	13.0			$\pm 3.0$	12.5	12.5	12.5			$\pm 3.0$
PLAQ.	PLTS	$10^9/mm^3 ; 10^9/l$	71	71	78			$\pm 20$	242	242	245			$\pm 40$	495	495	500			$\pm 60$
VPM	MPV	$\mu m^3 ; fl$	8.2	8.2	8.0			$\pm 2.0$	7.5	7.5	7.5			$\pm 2.0$	7.3	7.3	7.3			$\pm 2.0$
LYMPHO		#	1.2	1.2	1.3			$\pm 0.2$	2.4	2.4	2.5			$\pm 0.5$	3.0	3.0	3.0			$\pm 1.0$
		%	64.0	64.0	63.5			$\pm 8.0$	32.0	32.0	33.0			$\pm 6.0$	15.0	15.0	15.0			$\pm 5.0$
MONO		#	0.2	0.2	0.2			$\pm 0.2$	0.6	0.6	0.5			$\pm 0.4$	1.1	1.1	1.0			$\pm 1.0$
		%	12.0	12.0	11.5			$\pm 7.0$	7.5	7.5	7.0			$\pm 5.0$	5.5	5.5	5.0			$\pm 5.0$
GRANULO		#	0.5	0.5	0.5			$\pm 0.2$	4.5	4.5	4.5			$\pm 0.5$	15.7	15.7	15.9			$\pm 1.4$
		%	24.0	24.0	25.0			$\pm 7.0$	60.5	60.5	60.0			$\pm 6.0$	79.5	79.5	80.0			$\pm 7.0$
CRP		mg/dl	0.74	0.55	N/A			$\pm 0.30$	2.67	2.43	N/A			$\pm 0.70$	4.42	4.22	N/A			$\pm 1.00$
		mg/l	7.4	5.5	N/A			$\pm 3.0$	26.7	24.3	N/A			$\pm 7.0$	44.2	42.2	N/A			$\pm 10.0$

9930045-D

Ref: TEMP-1070 Rev.15