

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT

MC 103

CONTROL



(Exp.) 2014-01-05

(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | | UNITES UNITS | | ABX Lyse | | | | | | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|------------------------|-----------------|--------------|--------------|---------|-----------------|--------------|--------------|--------------|-----------------|---|-------------|--------------|--------------|-----------------|---|-------------------------|-------------|
| | | | | CONTROL | | | 1 | CONTROL | | | 2 | | CONTROL | | | 3 | | |
| | | | | MICROS | | | | MICROS | | | | | MICROS | | | | | |
| | | | | CRP | CRP 200 | 60 8P 60 18P | | CRP | CRP 200 | 60 8P 60 18P | | | CRP | CRP 200 | 60 8P 60 18P | | | |
| GB WBC | $10^3/mm^3; 10^9/l$ | 2.0 | 2.0 | 2.0 | | ± 0.4 | 7.4 | 7.4 | 7.6 | | | ± 1.0 | 19.8 | 19.8 | 20.3 | | | ± 2.2 |
| GR RBC | $10^6/mm^3; 10^{12}/l$ | 2.39 | 2.39 | 2.39 | | ± 0.15 | 4.57 | 4.57 | 4.59 | | | ± 0.18 | 5.53 | 5.53 | 5.59 | | | ± 0.25 |
| HB HGB | g/dl | 6.1 | 6.1 | 6.1 | | ± 0.4 | 13.4 | 13.4 | 13.4 | | | ± 0.5 | 17.5 | 17.5 | 17.5 | | | ± 0.7 |
| | g/l | 61 | 61 | 61 | | ± 4 | 134 | 134 | 134 | | | ± 5 | 175 | 175 | 175 | | | ± 7 |
| | mmol/l | 3.79 | 3.79 | 3.79 | | ± 0.25 | 8.32 | 8.32 | 8.32 | | | ± 0.31 | 10.87 | 10.87 | 10.87 | | | ± 0.44 |
| HT HCT | % | 16.7 | 16.7 | 16.7 | | ± 2.0 | 37.0 | 37.0 | 37.2 | | | ± 2.5 | 49.2 | 49.2 | 49.7 | | | ± 3.0 |
| | l/l | 0.167 | 0.167 | 0.167 | | ± 0.020 | 0.370 | 0.370 | 0.372 | | | ± 0.025 | 0.492 | 0.492 | 0.497 | | | ± 0.030 |
| VGM MCV | $\mu m^3; fl$ | 70 | 70 | 70 | | ± 4 | 81 | 81 | 81 | | | ± 4 | 89 | 89 | 89 | | | ± 4 |
| TGMH MCH | pg | 25.5 | 25.5 | 25.5 | | ± 2.4 | 29.3 | 29.3 | 29.2 | | | ± 2.4 | 31.6 | 31.6 | 31.3 | | | ± 2.8 |
| | fmol | 1.58 | 1.58 | 1.58 | | ± 0.15 | 1.82 | 1.82 | 1.81 | | | ± 0.15 | 1.97 | 1.97 | 1.94 | | | ± 0.18 |
| CCMH MCHC | g/dl | 36.5 | 36.5 | 36.5 | | ± 3.0 | 36.2 | 36.2 | 36.0 | | | ± 3.0 | 35.6 | 35.6 | 35.2 | | | ± 3.0 |
| | g/l | 365 | 365 | 365 | | ± 30 | 362 | 362 | 360 | | | ± 30 | 356 | 356 | 352 | | | ± 30 |
| | mmol/l | 22.64 | 22.64 | 22.64 | | ± 1.86 | 22.48 | 22.48 | 22.38 | | | ± 1.86 | 22.08 | 22.08 | 21.84 | | | ± 1.86 |
| IDR RDW | % | 14.0 | 14.0 | 14.0 | | ± 3.0 | 13.5 | 13.5 | 13.0 | | | ± 3.0 | 12.5 | 12.5 | 12.5 | | | ± 3.0 |
| PLAQ. PLTS | $10^3/mm^3; 10^9/l$ | 75 | 75 | 75 | | ± 20 | 240 | 240 | 246 | | | ± 40 | 490 | 490 | 495 | | | ± 60 |
| VPM MPV | $\mu m^3; fl$ | 8.7 | 8.7 | 8.1 | | ± 2.0 | 7.7 | 7.7 | 7.3 | | | ± 2.0 | 7.6 | 7.6 | 7.3 | | | ± 2.0 |
| LYMPHO | # | 1.3 | 1.3 | 1.3 | | ± 0.2 | 2.4 | 2.4 | 2.5 | | | ± 0.5 | 3.0 | 3.0 | 3.0 | | | ± 1.2 |
| | % | 64.0 | 64.0 | 62.5 | | ± 8.0 | 32.5 | 32.5 | 32.5 | | | ± 6.0 | 15.0 | 15.0 | 15.0 | | | ± 6.0 |
| MONO | # | 0.2 | 0.2 | 0.2 | | ± 0.2 | 0.7 | 0.7 | 0.6 | | | ± 0.4 | 1.4 | 1.4 | 1.2 | | | ± 1.0 |
| | % | 10.0 | 10.0 | 10.5 | | ± 7.0 | 9.5 | 9.5 | 8.5 | | | ± 5.0 | 7.0 | 7.0 | 6.0 | | | ± 5.0 |
| GRANULO | # | 0.5 | 0.5 | 0.5 | | ± 0.2 | 4.3 | 4.3 | 4.5 | | | ± 0.5 | 15.4 | 15.4 | 16.0 | | | ± 1.5 |
| | % | 26.0 | 26.0 | 27.0 | | ± 7.0 | 58.0 | 58.0 | 59.0 | | | ± 6.0 | 78.0 | 78.0 | 79.0 | | | ± 7.0 |
| CRP | mg/dl | 0.84 | 0.50 | N/A | | ± 0.30 | 2.52 | 1.93 | N/A | | | ± 0.70 | 4.44 | 3.71 | N/A | | | ± 1.00 |
| | mg/l | 8.4 | 5.0 | N/A | | ± 3.0 | 25.2 | 19.3 | N/A | | | ± 7.0 | 44.4 | 37.1 | N/A | | | ± 10.0 |

9930074-A FRONT/RECTO Ref: TEMP-1070 Rev.31