

ABX Minotrol CRP



LOT

MC 410

CONTROL



(Exp.) 2019-04-05
(YYYY - MM - DD)

Rev 1

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.6	7.7			± 1.0	19.9	20.4			± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.37	2.36			± 0.15	4.75	4.74			± 0.18	5.73	5.70			± 0.25			
HB HGB	g/dl	6.1	6.1			± 0.4	13.5	13.6			± 0.5	18.0	18.0			± 0.7			
	g/l	61	61			± 4	135	136			± 5	180	180			± 7			
HT HCT	mmol/l	3.79	3.79			± 0.25	8.38	8.45			± 0.31	11.18	11.18			± 0.44			
	%	16.6	16.5			± 2.0	38.0	37.9			± 2.5	51.0	50.7			± 3.0			
	l/l	0.166	0.165			± 0.020	0.380	0.379			± 0.025	0.510	0.507			± 0.030			
VGM MCV	µm ³ ; fl	70	70			± 4	80	80			± 4	89	89			± 4			
TGMH MCH	pg	25.7	25.8			± 2.4	28.4	28.7			± 2.4	31.4	31.6			± 2.8			
	fmol	1.60	1.61			± 0.15	1.76	1.78			± 0.15	1.95	1.96			± 0.18			
CCMH MCHC	g/dl	36.8	36.9			± 3.0	35.5	35.9			± 3.0	35.3	35.5			± 3.0			
	g/l	368	369			± 30	355	359			± 30	353	355			± 30			
	mmol/l	22.83	22.93			± 1.86	22.06	22.27			± 1.86	21.92	22.03			± 1.86			
IDR RDW	%	13.0	13.0			± 3.0	13.5	13.0			± 3.0	13.0	12.5			± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	73			± 20	265	265			± 40	515	516			± 60			
VPM MPV	µm ³ ; fl	8.4	7.6			± 2.0	8.4	7.7			± 2.0	8.0	7.5			± 2.0			
LYMPHO	#	1.3	1.3			± 0.2	2.6	2.6			± 0.5	3.0	3.1			± 1.2			
	%	65.0	63.0			± 8.0	34.0	34.0			± 6.0	15.0	15.0			± 6.0			
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.2	1.0			± 1.0			
	%	9.5	10.0			± 7.0	8.0	7.0			± 5.0	6.0	5.0			± 5.0			
GRANULO	#	0.5	0.5			± 0.2	4.4	4.5			± 0.5	15.7	16.3			± 1.4			
	%	25.5	27.0			± 7.0	58.0	59.0			± 6.0	79.0	80.0			± 7.0			
CRP	mg/dl	0.52	N/A			± 0.30	1.95	N/A			± 0.70	3.57	N/A			± 1.00			
	mg/l	5.2	N/A			± 3.0	19.5	N/A			± 7.0	35.7	N/A			± 10.0			

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514