

# ABX Minotrol CRP



**LOT**

**MC 401**

**CONTROL**



**(Exp.) 2017-01-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse													
				CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE	
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>	<b>2.1</b>	± 0.4	<b>7.5</b>	<b>7.5</b>	<b>7.6</b>	± 1.0	<b>20.0</b>	<b>20.0</b>	<b>20.2</b>	± 2.2				
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.43</b>	<b>2.43</b>	<b>2.40</b>	± 0.15	<b>4.60</b>	<b>4.60</b>	<b>4.60</b>	± 0.18	<b>5.67</b>	<b>5.67</b>	<b>5.70</b>	± 0.25				
HB HGB	g/dl	<b>6.3</b>	<b>6.3</b>	<b>6.2</b>	± 0.4	<b>13.6</b>	<b>13.6</b>	<b>13.5</b>	± 0.5	<b>18.0</b>	<b>18.0</b>	<b>17.9</b>	± 0.7				
	g/l	<b>63</b>	<b>63</b>	<b>62</b>	± 4	<b>136</b>	<b>136</b>	<b>135</b>	± 5	<b>180</b>	<b>180</b>	<b>179</b>	± 7				
HT HCT	mmol/l	<b>3.91</b>	<b>3.91</b>	<b>3.85</b>	± 0.25	<b>8.45</b>	<b>8.45</b>	<b>8.38</b>	± 0.31	<b>11.18</b>	<b>11.18</b>	<b>11.12</b>	± 0.44				
	%	<b>17.0</b>	<b>17.0</b>	<b>16.8</b>	± 2.0	<b>37.7</b>	<b>37.7</b>	<b>37.3</b>	± 2.5	<b>49.9</b>	<b>49.9</b>	<b>50.2</b>	± 3.0				
VGM MCV	l/l	<b>0.170</b>	<b>0.170</b>	<b>0.168</b>	± 0.020	<b>0.377</b>	<b>0.377</b>	<b>0.373</b>	± 0.025	<b>0.499</b>	<b>0.499</b>	<b>0.502</b>	± 0.030				
	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>	<b>70</b>	± 4	<b>82</b>	<b>82</b>	<b>81</b>	± 4	<b>88</b>	<b>88</b>	<b>88</b>	± 4				
TGMH MCH	pg	<b>25.9</b>	<b>25.9</b>	<b>25.8</b>	± 2.4	<b>29.6</b>	<b>29.6</b>	<b>29.3</b>	± 2.4	<b>31.7</b>	<b>31.7</b>	<b>31.4</b>	± 2.8				
	fmol	<b>1.61</b>	<b>1.61</b>	<b>1.60</b>	± 0.15	<b>1.84</b>	<b>1.84</b>	<b>1.82</b>	± 0.15	<b>1.97</b>	<b>1.97</b>	<b>1.95</b>	± 0.18				
CCMH MCHC	g/dl	<b>37.0</b>	<b>37.0</b>	<b>36.9</b>	± 3.0	<b>36.1</b>	<b>36.1</b>	<b>36.2</b>	± 3.0	<b>36.1</b>	<b>36.1</b>	<b>35.7</b>	± 3.0				
	g/l	<b>370</b>	<b>370</b>	<b>369</b>	± 30	<b>361</b>	<b>361</b>	<b>362</b>	± 30	<b>361</b>	<b>361</b>	<b>357</b>	± 30				
	mmol/l	<b>23.00</b>	<b>23.00</b>	<b>22.92</b>	± 1.86	<b>22.39</b>	<b>22.39</b>	<b>22.50</b>	± 1.86	<b>22.40</b>	<b>22.40</b>	<b>22.16</b>	± 1.86				
IDR RDW	%	<b>14.0</b>	<b>14.0</b>	<b>13.5</b>	± 3.0	<b>14.0</b>	<b>14.0</b>	<b>14.0</b>	± 3.0	<b>13.0</b>	<b>13.0</b>	<b>13.0</b>	± 3.0				
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>75</b>	<b>75</b>	<b>74</b>	± 20	<b>246</b>	<b>246</b>	<b>247</b>	± 40	<b>480</b>	<b>480</b>	<b>484</b>	± 60				
VPM MPV	µm <sup>3</sup> ; fl	<b>8.0</b>	<b>8.0</b>	<b>7.0</b>	± 2.0	<b>7.8</b>	<b>7.8</b>	<b>7.0</b>	± 2.0	<b>7.5</b>	<b>7.5</b>	<b>7.0</b>	± 2.0				
LYMPHO	#	<b>1.3</b>	<b>1.3</b>	<b>1.3</b>	± 0.3	<b>2.4</b>	<b>2.4</b>	<b>2.4</b>	± 0.5	<b>3.1</b>	<b>3.1</b>	<b>3.1</b>	± 1.3				
	%	<b>63.0</b>	<b>63.0</b>	<b>62.0</b>	± 8.0	<b>32.5</b>	<b>32.5</b>	<b>32.0</b>	± 6.0	<b>15.5</b>	<b>15.5</b>	<b>15.5</b>	± 6.0				
MONO	#	<b>0.3</b>	<b>0.3</b>	<b>0.3</b>	± 0.2	<b>0.6</b>	<b>0.6</b>	<b>0.5</b>	± 0.4	<b>1.1</b>	<b>1.1</b>	<b>1.1</b>	± 1.0				
	%	<b>12.0</b>	<b>12.0</b>	<b>12.0</b>	± 7.0	<b>7.5</b>	<b>7.5</b>	<b>7.0</b>	± 5.0	<b>5.5</b>	<b>5.5</b>	<b>5.5</b>	± 5.0				
GRANULO	#	<b>0.5</b>	<b>0.5</b>	<b>0.5</b>	± 0.3	<b>4.5</b>	<b>4.5</b>	<b>4.6</b>	± 0.5	<b>15.8</b>	<b>15.8</b>	<b>16.0</b>	± 1.5				
	%	<b>25.0</b>	<b>25.0</b>	<b>26.0</b>	± 8.0	<b>60.0</b>	<b>60.0</b>	<b>61.0</b>	± 6.0	<b>79.0</b>	<b>79.0</b>	<b>79.0</b>	± 7.0				
CRP	mg/dl	<b>0.65</b>	<b>0.48</b>	N/A	± 0.30	<b>2.40</b>	<b>2.36</b>	N/A	± 0.70	<b>4.09</b>	<b>4.20</b>	N/A	± 1.00				
	mg/l	<b>6.5</b>	<b>4.8</b>	N/A	± 3.0	<b>24.0</b>	<b>23.6</b>	N/A	± 7.0	<b>40.9</b>	<b>42.0</b>	N/A	± 10.0				

Ref: TEMP-1070 Rev.36 FRONT/RECTO 9930089-A