

ABX Minotrol CRP



LOT

MC 421

CONTROL



(Exp.) 2022-01-05
(YYYY - MM - DD)

Rev 1

		ABX Lyse															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2	2.2			± 0.4	7.3	7.7			± 1.0	19.2	20.0			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.29	2.30			± 0.15	4.46	4.53			± 0.18	5.58	5.64			± 0.25	
HB HGB	g/dl	6.0	6.1			± 0.4	13.3	13.5			± 0.5	17.8	18.1			± 0.7	
	g/l	60	61			± 4	133	135			± 5	178	181			± 7	
HT HCT	mmol/l	3.73	3.79			± 0.25	8.26	8.38			± 0.31	11.05	11.24			± 0.44	
	%	16.5	16.6			± 2.0	37.0	37.1			± 2.5	50.8	51.3			± 3.0	
VGM MCV	l/l	0.165	0.166			± 0.020	0.370	0.371			± 0.025	0.508	0.513			± 0.030	
	µm ³ ; fl	72	72			± 4	83	82			± 4	91	91			± 4	
TGMH MCH	pg	26.2	26.5			± 2.4	29.8	29.8			± 2.4	31.9	32.1			± 2.8	
	fmol	1.63	1.65			± 0.15	1.85	1.85			± 0.15	1.98	1.99			± 0.18	
CCMH MCHC	g/dl	36.4	36.8			± 3.0	35.9	36.3			± 3.0	35.1	35.3			± 3.0	
	g/l	364	368			± 30	359	363			± 30	351	353			± 30	
IDR RDW	mmol/l	22.60	22.88			± 1.86	22.31	22.57			± 1.86	21.77	21.90			± 1.86	
	%	13.5	13.5			± 3.0	14.0	13.5			± 3.0	13.5	13.0			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	74	74			± 20	256	256			± 40	510	505			± 60	
VPM MPV	µm ³ ; fl	8.8	8.2			± 2.0	8.0	7.5			± 2.0	7.8	7.4			± 2.0	
LYMPHO	#	1.4	1.4			± 0.2	2.4	2.5			± 0.5	3.1	2.9			± 1.1	
	%	64.5	62.5			± 8.0	33.5	32.0			± 6.0	16.0	14.5			± 5.0	
MONO	#	0.2	0.2			± 0.2	0.6	0.6			± 0.4	1.0	1.0			± 1.0	
	%	10.0	9.5			± 7.0	8.5	8.0			± 5.0	5.0	5.0			± 5.0	
GRANULO	#	0.6	0.6			± 0.2	4.2	4.6			± 0.5	15.2	16.1			± 1.5	
	%	25.5	28.0			± 7.0	58.0	60.0			± 6.0	79.0	80.5			± 7.0	
CRP	mg/dl	0.56	N/A			± 0.30	2.15	N/A			± 0.70	3.40	N/A			± 1.00	
	mg/l	5.6	N/A			± 3.0	21.5	N/A			± 7.0	34.0	N/A			± 10.0	

Ref: TEMP-1070 Rev.39 FRONT/RECTO 1300092170