

# ABX Minotrol CRP



**LOT**

**MC 403**

**CONTROL**



**(Exp.) 2017-07-05**  
(YYYY - MM - DD)

Rev 1

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				MICROS		MICROS				MICROS		MICROS					
		CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	2.0	2.0	2.0		± 0.4	7.7	7.7	7.7		± 1.0	20.3	20.3	20.6		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	2.25	2.25	2.27		± 0.15	4.68	4.68	4.70		± 0.18	5.70	5.70	5.78		± 0.25			
HB HGB	g/dl	5.9	5.9	5.9		± 0.4	13.5	13.5	13.5		± 0.5	17.8	17.8	17.8		± 0.7			
	g/l	59	59	59		± 4	135	135	135		± 5	178	178	178		± 7			
HT HCT	mmol/l	3.66	3.66	3.66		± 0.25	8.38	8.38	8.38		± 0.31	11.05	11.05	11.05		± 0.44			
	%	16.0	16.0	15.9		± 2.0	37.9	37.9	37.6		± 2.5	50.7	50.7	50.9		± 3.0			
VGM MCV	l/l	0.160	0.160	0.159		± 0.020	0.379	0.379	0.376		± 0.025	0.507	0.507	0.509		± 0.030			
	µm <sup>3</sup> ; fl	71	71	70		± 4	81	81	80		± 4	89	89	88		± 4			
TGMH MCH	pg	26.2	26.2	26.0		± 2.4	28.8	28.8	28.7		± 2.4	31.2	31.2	30.8		± 2.8			
	fmol	1.63	1.63	1.61		± 0.15	1.79	1.79	1.78		± 0.15	1.94	1.94	1.91		± 0.18			
CCMH MCHC	g/dl	36.9	36.9	37.1		± 3.0	35.6	35.6	35.9		± 3.0	35.1	35.1	35.0		± 3.0			
	g/l	369	369	371		± 30	356	356	359		± 30	351	351	350		± 30			
	mmol/l	22.94	22.94	23.06		± 1.86	22.12	22.12	22.30		± 1.86	21.79	21.79	21.73		± 1.86			
IDR RDW	%	14.0	14.0	12.5		± 3.0	14.0	14.0	12.5		± 3.0	13.5	13.5	12.0		± 3.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	75	75	75		± 20	265	265	266		± 40	515	515	520		± 60			
VPM MPV	µm <sup>3</sup> ; fl	9.2	9.2	8.6		± 2.0	8.7	8.7	8.2		± 2.0	8.4	8.4	7.9		± 2.0			
LYMPHO	#	1.3	1.3	1.2		± 0.3	2.7	2.7	2.8		± 0.6	3.2	3.2	3.4		± 1.3			
	%	62.5	62.5	62.0		± 7.0	34.5	34.5	36.0		± 6.0	16.0	16.0	16.5		± 6.0			
MONO	#	0.2	0.2	0.2		± 0.2	0.7	0.7	0.5		± 0.4	1.2	1.2	1.0		± 1.0			
	%	9.5	9.5	9.0		± 6.0	8.5	8.5	7.0		± 5.0	6.0	6.0	5.0		± 5.0			
GRANULO	#	0.6	0.6	0.6		± 0.3	4.4	4.4	4.4		± 0.6	15.8	15.8	16.2		± 1.5			
	%	28.0	28.0	29.0		± 8.0	57.0	57.0	57.0		± 6.0	78.0	78.0	78.5		± 7.0			
CRP	mg/dl	0.95	0.76	N/A		± 0.30	2.73	2.49	N/A		± 0.70	4.35	3.94	N/A		± 1.00			
	mg/l	9.5	7.6	N/A		± 3.0	27.3	24.9	N/A		± 7.0	43.5	39.4	N/A		± 10.0			

Ref: TEMP-1070 Rev.37 FRONT/RECTO 9930089-A