

ABX Minotrol CRP



LOT

MC 423

CONTROL



(Exp.) **2022-07-05**
(YYYY - MM - DD)

Rev 2

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1			± 0.4	7.5	7.5			± 1.0	20.4	20.8			± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.22	2.19			± 0.15	4.53	4.53			± 0.18	5.67	5.76			± 0.25
HB HGB	g/dl	5.9	5.9			± 0.4	13.3	13.4			± 0.5	18.0	18.2			± 0.7
	g/l	59	59			± 4	133	134			± 5	180	182			± 7
HT HCT	mmol/l	3.66	3.66			± 0.25	8.26	8.32			± 0.31	11.18	11.30			± 0.44
	%	16.0	16.0			± 2.0	36.7	36.7			± 2.5	51.0	51.8			± 3.0
VGM MCV	l/l	0.160	0.160			± 0.020	0.367	0.367			± 0.025	0.510	0.518			± 0.030
	µm ³ ; fl	72	73			± 4	81	81			± 4	90	90			± 4
TGMH MCH	pg	26.6	26.9			± 2.4	29.4	29.6			± 2.4	31.7	31.6			± 2.8
	fmol	1.65	1.67			± 0.15	1.82	1.84			± 0.15	1.97	1.96			± 0.18
CCMH MCHC	g/dl	36.9	36.9			± 3.0	36.2	36.5			± 3.0	35.3	35.1			± 3.0
	g/l	369	369			± 30	362	365			± 30	353	351			± 30
	mmol/l	22.92	22.92			± 1.86	22.51	22.68			± 1.86	21.90	21.80			± 1.86
IDR RDW	%	13.5	13.5			± 3.0	13.0	12.5			± 3.0	12.5	12.0			± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	80	80			± 20	253	255			± 40	495	500			± 60
VPM MPV	µm ³ ; fl	8.2	7.8			± 2.0	8.1	7.8			± 2.0	7.9	7.7			± 2.0
LYMPHO	#	1.2	1.2			± 0.2	2.5	2.4			± 0.5	3.6	3.5			± 1.1
	%	59.5	58.5			± 8.0	33.0	32.5			± 6.0	17.5	17.0			± 5.0
MONO	#	0.3	0.3			± 0.2	0.7	0.6			± 0.4	1.4	1.2			± 1.0
	%	13.5	12.5			± 7.0	9.5	8.5			± 5.0	7.0	6.0			± 5.0
GRANULO	#	0.6	0.6			± 0.2	4.3	4.4			± 0.5	15.4	16.0			± 1.5
	%	27.0	29.0			± 7.0	57.5	59.0			± 6.0	75.5	77.0			± 7.0
CRP	mg/dl	0.42	N/A			± 0.30	2.66	N/A			± 0.70	4.16	N/A			± 1.00
	mg/l	4.2	N/A			± 3.0	26.6	N/A			± 7.0	41.6	N/A			± 10.0

Ref: TEMP-1070 Rev.39 FRONT/RECTO 1300092170