

# ABX Minotrol CRP



**LOT**

**MC 408**

**CONTROL**



**(Exp.) 2018-10-05**  
(YYYY - MM - DD)

Rev 1

		<b>ABX Lyse</b>																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				MICROS		MICROS				MICROS		MICROS					
		CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>	<b>2.1</b>		± 0.4	<b>7.5</b>	<b>7.5</b>	<b>7.4</b>		± 1.0	<b>20.4</b>	<b>20.4</b>	<b>20.4</b>		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.29</b>	<b>2.29</b>	<b>2.29</b>		± 0.15	<b>4.60</b>	<b>4.60</b>	<b>4.58</b>		± 0.18	<b>5.55</b>	<b>5.55</b>	<b>5.60</b>		± 0.25			
HB HGB	g/dl	<b>6.0</b>	<b>6.0</b>	<b>6.0</b>		± 0.4	<b>13.3</b>	<b>13.3</b>	<b>13.4</b>		± 0.5	<b>17.8</b>	<b>17.8</b>	<b>17.9</b>		± 0.7			
	g/l	<b>60</b>	<b>60</b>	<b>60</b>		± 4	<b>133</b>	<b>133</b>	<b>134</b>		± 5	<b>178</b>	<b>178</b>	<b>179</b>		± 7			
HT HCT	mmol/l	<b>3.73</b>	<b>3.73</b>	<b>3.73</b>		± 0.25	<b>8.26</b>	<b>8.26</b>	<b>8.32</b>		± 0.31	<b>11.05</b>	<b>11.05</b>	<b>11.12</b>		± 0.44			
	%	<b>16.3</b>	<b>16.3</b>	<b>16.3</b>		± 2.0	<b>37.3</b>	<b>37.3</b>	<b>37.1</b>		± 2.5	<b>49.4</b>	<b>49.4</b>	<b>50.4</b>		± 3.0			
VGM MCV	l/l	<b>0.163</b>	<b>0.163</b>	<b>0.163</b>		± 0.020	<b>0.373</b>	<b>0.373</b>	<b>0.371</b>		± 0.025	<b>0.494</b>	<b>0.494</b>	<b>0.504</b>		± 0.030			
	µm <sup>3</sup> ; fl	<b>71</b>	<b>71</b>	<b>71</b>		± 4	<b>81</b>	<b>81</b>	<b>81</b>		± 4	<b>89</b>	<b>89</b>	<b>90</b>		± 4			
TGMH MCH	pg	<b>26.2</b>	<b>26.2</b>	<b>26.2</b>		± 2.4	<b>28.9</b>	<b>28.9</b>	<b>29.3</b>		± 2.4	<b>32.1</b>	<b>32.1</b>	<b>32.0</b>		± 2.8			
	fmol	<b>1.63</b>	<b>1.63</b>	<b>1.63</b>		± 0.15	<b>1.80</b>	<b>1.80</b>	<b>1.82</b>		± 0.15	<b>1.99</b>	<b>1.99</b>	<b>1.98</b>		± 0.18			
CCMH MCHC	g/dl	<b>36.9</b>	<b>36.9</b>	<b>36.9</b>		± 3.0	<b>35.7</b>	<b>35.7</b>	<b>36.1</b>		± 3.0	<b>36.0</b>	<b>36.0</b>	<b>35.5</b>		± 3.0			
	g/l	<b>369</b>	<b>369</b>	<b>369</b>		± 30	<b>357</b>	<b>357</b>	<b>361</b>		± 30	<b>360</b>	<b>360</b>	<b>355</b>		± 30			
	mmol/l	<b>22.92</b>	<b>22.92</b>	<b>22.92</b>		± 1.86	<b>22.17</b>	<b>22.17</b>	<b>22.43</b>		± 1.86	<b>22.38</b>	<b>22.38</b>	<b>22.06</b>		± 1.86			
IDR RDW	%	<b>13.0</b>	<b>13.0</b>	<b>13.2</b>		± 3.0	<b>13.5</b>	<b>13.5</b>	<b>13.5</b>		± 3.0	<b>12.4</b>	<b>12.4</b>	<b>11.8</b>		± 3.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>76</b>	<b>76</b>	<b>79</b>		± 20	<b>265</b>	<b>265</b>	<b>265</b>		± 40	<b>525</b>	<b>525</b>	<b>525</b>		± 60			
VPM MPV	µm <sup>3</sup> ; fl	<b>8.4</b>	<b>8.4</b>	<b>7.5</b>		± 2.0	<b>7.7</b>	<b>7.7</b>	<b>7.1</b>		± 2.0	<b>7.8</b>	<b>7.8</b>	<b>7.3</b>		± 2.0			
LYMPHO	#	<b>1.4</b>	<b>1.4</b>	<b>1.3</b>		± 0.2	<b>2.6</b>	<b>2.6</b>	<b>2.5</b>		± 0.5	<b>3.0</b>	<b>3.0</b>	<b>3.0</b>		± 1.3			
	%	<b>66.0</b>	<b>66.0</b>	<b>63.5</b>		± 8.0	<b>34.0</b>	<b>34.0</b>	<b>33.5</b>		± 6.0	<b>14.5</b>	<b>14.5</b>	<b>14.5</b>		± 6.0			
MONO	#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>		± 0.2	<b>0.5</b>	<b>0.5</b>	<b>0.5</b>		± 0.4	<b>1.2</b>	<b>1.2</b>	<b>1.1</b>		± 1.0			
	%	<b>8.0</b>	<b>8.0</b>	<b>9.5</b>		± 7.0	<b>7.0</b>	<b>7.0</b>	<b>7.0</b>		± 5.0	<b>6.0</b>	<b>6.0</b>	<b>5.5</b>		± 5.0			
GRANULO	#	<b>0.5</b>	<b>0.5</b>	<b>0.6</b>		± 0.2	<b>4.4</b>	<b>4.4</b>	<b>4.4</b>		± 0.5	<b>16.2</b>	<b>16.2</b>	<b>16.3</b>		± 1.5			
	%	<b>26.0</b>	<b>26.0</b>	<b>27.0</b>		± 7.0	<b>59.0</b>	<b>59.0</b>	<b>59.5</b>		± 6.0	<b>79.5</b>	<b>79.5</b>	<b>80.0</b>		± 7.0			
CRP	mg/dl	<b>0.68</b>	<b>0.49</b>	N/A		± 0.30	<b>2.56</b>	<b>2.28</b>	N/A		± 0.70	<b>5.07</b>	<b>4.38</b>	N/A		± 1.00			
	mg/l	<b>6.8</b>	<b>4.9</b>	N/A		± 3.0	<b>25.6</b>	<b>22.8</b>	N/A		± 7.0	<b>50.7</b>	<b>43.8</b>	N/A		± 10.0			

Ref: TEMP-1070 Rev.37 FRONT/RECTO 9930089-A