

ABX Minotrol CRP



LOT

MC 414

CONTROL



(Exp.) **2020-04-05**
(YYYY-MM-DD)

Rev 1

		ABX Lyse																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					1	CONTROL					2	CONTROL					3	TOLERANCES TOLERANCE
		MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	1.8	2.0				± 0.4	7.3	7.8				± 1.0	19.4	20.5				± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.40	2.41				± 0.15	4.55	4.59				± 0.18	5.60	5.65				± 0.25	
HB HGB	g/dl	6.1	6.1				± 0.4	13.3	13.4				± 0.5	17.5	17.7				± 0.7	
	g/l	61	61				± 4	133	134				± 5	175	177				± 7	
	mmol/l	3.79	3.79				± 0.25	8.26	8.32				± 0.31	10.87	10.99				± 0.44	
HT HCT	%	16.8	16.9				± 2.0	36.9	37.2				± 2.5	49.8	50.3				± 3.0	
	l/l	0.168	0.169				± 0.020	0.369	0.372				± 0.025	0.498	0.503				± 0.030	
VGM MCV	µm ³ ; fl	70	70				± 4	81	81				± 4	89	89				± 4	
TGMH MCH	pg	25.4	25.3				± 2.4	29.2	29.2				± 2.4	31.3	31.3				± 2.8	
	fmol	1.58	1.57				± 0.15	1.82	1.81				± 0.15	1.94	1.95				± 0.18	
CCMH MCHC	g/dl	36.3	36.2				± 3.0	36.1	36.0				± 3.0	35.1	35.2				± 3.0	
	g/l	363	362				± 30	361	360				± 30	351	352				± 30	
	mmol/l	22.55	22.45				± 1.86	22.41	22.38				± 1.86	21.80	21.86				± 1.86	
IDR RDW	%	14.5	14.0				± 3.0	14.0	13.5				± 3.0	13.5	13.0				± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	77	75				± 20	263	265				± 40	550	565				± 60	
VPM MPV	µm ³ ; fl	8.3	7.7				± 2.0	7.9	7.5				± 2.0	7.9	7.5				± 2.0	
LYMPHO	#	1.2	1.3				± 0.2	2.5	2.7				± 0.5	3.2	3.4				± 1.3	
	%	66.0	64.5				± 8.0	34.0	34.0				± 6.0	16.5	16.5				± 6.0	
MONO	#	0.2	0.2				± 0.2	0.5	0.5				± 0.4	1.2	0.9				± 0.8	
	%	9.5	9.0				± 7.0	7.0	6.5				± 5.0	6.0	4.5				± 4.0	
GRANULO	#	0.4	0.5				± 0.2	4.3	4.6				± 0.5	15.0	16.2				± 1.5	
	%	24.5	26.5				± 7.0	59.0	59.5				± 6.0	77.5	79.0				± 7.0	
CRP	mg/dl	0.50	N/A				± 0.30	2.62	N/A				± 0.70	4.09	N/A				± 1.00	
	mg/l	5.0	N/A				± 3.0	26.2	N/A				± 7.0	40.9	N/A				± 10.0	

1300060514 FRONT/RECTO Ref: TEMP-1070 Rev.38