

ABX Minotrol CRP



LOT

MC 415

CONTROL



(Exp.) **2020-07-05**
(YYYY - MM - DD)

Rev 1

| | | ABX Lyse | | | | | | | | | | | | | | | |
|--------------------------|---|----------|-----------------|--|--|-------------------------|---------|-----------------|--|--|-------------------------|---------|-----------------|--|--|-------------------------|-------------------------|
| PARAMETRES PARAMETERS | UNITES UNITS | CONTROL | | | | 1 | CONTROL | | | | 2 | CONTROL | | | | 3 | TOLERANCES TOLERANCE |
| | | MICROS | | | | TOLERANCES TOLERANCE | MICROS | | | | TOLERANCES TOLERANCE | MICROS | | | | TOLERANCES TOLERANCE | |
| | | CRP 200 | 60 8P 60 18P | | | | CRP 200 | 60 8P 60 18P | | | | CRP 200 | 60 8P 60 18P | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.0 | 2.0 | | | ± 0.4 | 7.5 | 7.6 | | | ± 1.0 | 20.4 | 20.8 | | | ± 2.2 | |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.43 | 2.43 | | | ± 0.15 | 4.61 | 4.68 | | | ± 0.18 | 5.62 | 5.68 | | | ± 0.25 | |
| HB HGB | g/dl | 6.2 | 6.2 | | | ± 0.4 | 13.3 | 13.5 | | | ± 0.5 | 18.0 | 18.3 | | | ± 0.7 | |
| | g/l | 62 | 62 | | | ± 4 | 133 | 135 | | | ± 5 | 180 | 183 | | | ± 7 | |
| | mmol/l | 3.85 | 3.85 | | | ± 0.25 | 8.26 | 8.38 | | | ± 0.31 | 11.18 | 11.36 | | | ± 0.44 | |
| HT HCT | % | 16.8 | 16.8 | | | ± 2.0 | 36.9 | 37.4 | | | ± 2.5 | 50.6 | 51.1 | | | ± 3.0 | |
| | l/l | 0.168 | 0.168 | | | ± 0.020 | 0.369 | 0.374 | | | ± 0.025 | 0.506 | 0.511 | | | ± 0.030 | |
| VGM MCV | µm ³ ; fl | 69 | 69 | | | ± 4 | 80 | 80 | | | ± 4 | 90 | 90 | | | ± 4 | |
| TGMH MCH | pg | 25.5 | 25.5 | | | ± 2.4 | 28.9 | 28.8 | | | ± 2.4 | 32.0 | 32.2 | | | ± 2.8 | |
| | fmol | 1.58 | 1.58 | | | ± 0.15 | 1.79 | 1.79 | | | ± 0.15 | 1.99 | 2.00 | | | ± 0.18 | |
| CCMH MCHC | g/dl | 37.0 | 37.0 | | | ± 3.0 | 36.1 | 36.1 | | | ± 3.0 | 35.6 | 35.8 | | | ± 3.0 | |
| | g/l | 370 | 370 | | | ± 30 | 361 | 361 | | | ± 30 | 356 | 358 | | | ± 30 | |
| | mmol/l | 22.96 | 22.96 | | | ± 1.86 | 22.40 | 22.39 | | | ± 1.86 | 22.10 | 22.23 | | | ± 1.86 | |
| IDR RDW | % | 13.7 | 13.6 | | | ± 3.0 | 14.6 | 13.8 | | | ± 3.0 | 13.5 | 13.0 | | | ± 3.0 | |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 65 | 70 | | | ± 20 | 245 | 255 | | | ± 40 | 500 | 510 | | | ± 60 | |
| VPM MPV | µm ³ ; fl | 8.8 | 7.9 | | | ± 2.0 | 8.1 | 7.4 | | | ± 2.0 | 8.0 | 7.4 | | | ± 2.0 | |
| LYMPHO | # | 1.3 | 1.2 | | | ± 0.2 | 2.6 | 2.6 | | | ± 0.5 | 3.5 | 3.4 | | | ± 1.3 | |
| | % | 64.5 | 62.0 | | | ± 8.0 | 34.0 | 34.5 | | | ± 6.0 | 17.0 | 16.5 | | | ± 6.0 | |
| MONO | # | 0.2 | 0.2 | | | ± 0.2 | 0.6 | 0.5 | | | ± 0.4 | 1.2 | 0.9 | | | ± 0.8 | |
| | % | 10.5 | 11.0 | | | ± 7.0 | 7.5 | 6.5 | | | ± 5.0 | 6.0 | 4.5 | | | ± 4.0 | |
| GRANULO | # | 0.5 | 0.5 | | | ± 0.2 | 4.4 | 4.5 | | | ± 0.5 | 15.7 | 16.4 | | | ± 1.5 | |
| | % | 25.0 | 27.0 | | | ± 7.0 | 58.5 | 59.0 | | | ± 6.0 | 77.0 | 79.0 | | | ± 7.0 | |
| CRP | mg/dl | 0.59 | N/A | | | ± 0.30 | 2.11 | N/A | | | ± 0.70 | 4.68 | N/A | | | ± 1.00 | |
| | mg/l | 5.9 | N/A | | | ± 3.0 | 21.1 | N/A | | | ± 7.0 | 46.8 | N/A | | | ± 10.0 | |

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514