

ABX Minotrol CRP



LOT

MC 416

CONTROL



(Exp.) **2020-10-05**
(YYYY - MM - DD)

Rev 2

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					1	CONTROL					2	CONTROL					3
		MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	1.9	2.0				± 0.4	7.6	7.8				± 1.0	20.0	20.5				± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.32	2.35				± 0.15	4.50	4.54				± 0.18	5.58	5.60				± 0.25
HB HGB	g/dl	6.0	6.1				± 0.4	13.2	13.4				± 0.5	17.7	17.9				± 0.7
	g/l	60	61				± 4	132	134				± 5	177	179				± 7
	mmol/l	3.73	3.79				± 0.25	8.20	8.32				± 0.31	10.99	11.12				± 0.44
HT HCT	%	16.2	16.5				± 2.0	36.9	36.8				± 2.5	49.7	49.8				± 3.0
	l/l	0.162	0.165				± 0.020	0.369	0.368				± 0.025	0.497	0.498				± 0.030
VGM MCV	µm ³ ; fl	70	70				± 4	82	81				± 4	89	89				± 4
TGMH MCH	pg	25.9	26.0				± 2.4	29.3	29.5				± 2.4	31.7	32.0				± 2.8
	fmol	1.61	1.61				± 0.15	1.82	1.83				± 0.15	1.97	1.98				± 0.18
CCMH MCHC	g/dl	36.9	37.1				± 3.0	35.8	36.4				± 3.0	35.6	35.9				± 3.0
	g/l	369	371				± 30	358	364				± 30	356	359				± 30
	mmol/l	22.94	23.03				± 1.86	22.21	22.63				± 1.86	22.13	22.30				± 1.86
IDR RDW	%	14.0	13.5				± 3.0	13.5	13.3				± 3.0	13.0	12.5				± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	68				± 20	265	270				± 40	505	509				± 60
VPM MPV	µm ³ ; fl	8.0	7.4				± 2.0	7.5	7.1				± 2.0	7.3	7.0				± 2.0
LYMPHO	#	1.2	1.2				± 0.2	2.6	2.6				± 0.5	3.4	3.3				± 1.3
	%	63.5	62.0				± 8.0	34.0	33.0				± 6.0	17.0	16.0				± 6.0
MONO	#	0.2	0.2				± 0.2	0.6	0.5				± 0.4	1.0	0.8				± 0.8
	%	10.5	10.5				± 7.0	7.5	7.0				± 5.0	5.0	4.0				± 4.0
GRANULO	#	0.5	0.6				± 0.2	4.4	4.7				± 0.5	15.6	16.4				± 1.5
	%	26.0	27.5				± 7.0	58.5	60.0				± 6.0	78.0	80.0				± 7.0
CRP	mg/dl	0.42	N/A				± 0.30	2.05	N/A				± 0.70	3.55	N/A				± 1.00
	mg/l	4.2	N/A				± 3.0	20.5	N/A				± 7.0	35.5	N/A				± 10.0

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514