

# ABX Minotrol CRP



**LOT**

**MC 417**

**CONTROL**



(Exp.) **2021-01-05**  
(YYYY - MM - DD)

Rev 1

		<b>ABX Lyse</b>																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>			± 0.4	<b>7.6</b>	<b>7.8</b>			± 1.0	<b>19.8</b>	<b>20.1</b>			± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.42</b>	<b>2.40</b>			± 0.15	<b>4.56</b>	<b>4.57</b>			± 0.18	<b>5.50</b>	<b>5.48</b>			± 0.25			
HB HGB	g/dl	<b>6.1</b>	<b>6.1</b>			± 0.4	<b>13.5</b>	<b>13.5</b>			± 0.5	<b>17.7</b>	<b>17.7</b>			± 0.7			
	g/l	<b>61</b>	<b>61</b>			± 4	<b>135</b>	<b>135</b>			± 5	<b>177</b>	<b>177</b>			± 7			
	mmol/l	<b>3.79</b>	<b>3.79</b>			± 0.25	<b>8.38</b>	<b>8.38</b>			± 0.31	<b>10.99</b>	<b>10.99</b>			± 0.44			
HT HCT	%	<b>16.9</b>	<b>17.0</b>			± 2.0	<b>37.8</b>	<b>37.9</b>			± 2.5	<b>50.1</b>	<b>49.9</b>			± 3.0			
	l/l	<b>0.169</b>	<b>0.170</b>			± 0.020	<b>0.378</b>	<b>0.379</b>			± 0.025	<b>0.501</b>	<b>0.499</b>			± 0.030			
VGM MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>71</b>			± 4	<b>83</b>	<b>83</b>			± 4	<b>91</b>	<b>91</b>			± 4			
TGMH MCH	pg	<b>25.2</b>	<b>25.4</b>			± 2.4	<b>29.6</b>	<b>29.5</b>			± 2.4	<b>32.2</b>	<b>32.3</b>			± 2.8			
	fmol	<b>1.57</b>	<b>1.58</b>			± 0.15	<b>1.84</b>	<b>1.83</b>			± 0.15	<b>2.00</b>	<b>2.01</b>			± 0.18			
CCMH MCHC	g/dl	<b>36.0</b>	<b>35.8</b>			± 3.0	<b>35.7</b>	<b>35.6</b>			± 3.0	<b>35.4</b>	<b>35.5</b>			± 3.0			
	g/l	<b>360</b>	<b>358</b>			± 30	<b>357</b>	<b>356</b>			± 30	<b>354</b>	<b>355</b>			± 30			
	mmol/l	<b>22.36</b>	<b>22.23</b>			± 1.86	<b>22.15</b>	<b>22.10</b>			± 1.86	<b>21.96</b>	<b>22.04</b>			± 1.86			
IDR RDW	%	<b>14.0</b>	<b>14.0</b>			± 3.0	<b>13.5</b>	<b>13.5</b>			± 3.0	<b>12.5</b>	<b>12.5</b>			± 3.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>74</b>	<b>75</b>			± 20	<b>252</b>	<b>257</b>			± 40	<b>507</b>	<b>503</b>			± 60			
VPM MPV	µm <sup>3</sup> ; fl	<b>8.9</b>	<b>7.9</b>			± 2.0	<b>7.9</b>	<b>7.3</b>			± 2.0	<b>7.3</b>	<b>6.8</b>			± 2.0			
LYMPHO	#	<b>1.3</b>	<b>1.3</b>			± 0.2	<b>2.7</b>	<b>2.8</b>			± 0.5	<b>3.4</b>	<b>3.3</b>			± 1.2			
	%	<b>66.0</b>	<b>65.0</b>			± 8.0	<b>35.5</b>	<b>36.0</b>			± 6.0	<b>17.0</b>	<b>16.5</b>			± 6.0			
MONO	#	<b>0.2</b>	<b>0.2</b>			± 0.2	<b>0.6</b>	<b>0.5</b>			± 0.4	<b>1.0</b>	<b>0.9</b>			± 0.9			
	%	<b>10.5</b>	<b>9.7</b>			± 7.0	<b>7.8</b>	<b>6.5</b>			± 5.0	<b>5.0</b>	<b>4.5</b>			± 4.5			
GRANULO	#	<b>0.5</b>	<b>0.5</b>			± 0.2	<b>4.3</b>	<b>4.5</b>			± 0.5	<b>15.4</b>	<b>15.9</b>			± 1.5			
	%	<b>23.5</b>	<b>25.3</b>			± 7.0	<b>56.7</b>	<b>57.5</b>			± 6.0	<b>78.0</b>	<b>79.0</b>			± 7.0			
CRP	mg/dl	<b>0.46</b>	N/A			± 0.30	<b>1.92</b>	N/A			± 0.70	<b>3.94</b>	N/A			± 1.00			
	mg/l	<b>4.6</b>	N/A			± 3.0	<b>19.2</b>	N/A			± 7.0	<b>39.4</b>	N/A			± 10.0			

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514