

ABX Minotrol CRP



LOT

MC 046

CONTROL



(Exp.) 2016-07-05
(YYYY - MM - DD)

Rev 1

		ABX Lyse																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					1	CONTROL					2	CONTROL					3	TOLERANCES TOLERANCE
		MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE	
		CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0	2.1			± 0.4	7.4	7.4	7.6			± 1.0	20.1	20.1	20.5			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.30	2.30	2.30			± 0.15	4.57	4.57	4.58			± 0.18	5.60	5.60	5.62			± 0.25	
HB HGB	g/dl	6.1	6.1	6.1			± 0.4	13.4	13.4	13.4			± 0.5	17.8	17.8	17.9			± 0.7	
	g/l	61	61	61			± 4	134	134	134			± 5	178	178	179			± 7	
	mmol/l	3.79	3.79	3.79			± 0.25	8.32	8.32	8.32			± 0.31	11.05	11.05	11.12			± 0.44	
HT HCT	%	16.3	16.3	16.3			± 2.0	37.0	37.0	37.1			± 2.5	49.8	49.8	50.0			± 3.0	
	l/l	0.163	0.163	0.163			± 0.020	0.370	0.370	0.371			± 0.025	0.498	0.498	0.500			± 0.030	
VGM MCV	µm ³ ; fl	71	71	71			± 4	81	81	81			± 4	89	89	89			± 4	
TGMH MCH	pg	26.5	26.5	26.5			± 2.4	29.3	29.3	29.3			± 2.4	31.8	31.8	31.9			± 2.8	
	fmol	1.65	1.65	1.65			± 0.15	1.82	1.82	1.82			± 0.15	1.97	1.97	1.98			± 0.18	
CCMH MCHC	g/dl	37.4	37.4	37.4			± 3.0	36.2	36.2	36.1			± 3.0	35.7	35.7	35.8			± 3.0	
	g/l	374	374	374			± 30	362	362	361			± 30	357	357	358			± 30	
	mmol/l	23.20	23.20	23.20			± 1.86	22.48	22.48	22.43			± 1.86	22.18	22.18	22.22			± 1.86	
IDR RDW	%	13.5	13.5	13.3			± 3.0	13.2	13.2	13.9			± 3.0	12.5	12.5	13.4			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	65	65	65			± 20	265	265	260			± 40	510	510	510			± 60	
VPM MPV	µm ³ ; fl	9.2	9.2	8.5			± 2.0	8.3	8.3	7.9			± 2.0	7.9	7.9	7.7			± 2.0	
LYMPHO	#	1.3	1.3	1.3			± 0.4	2.6	2.6	2.6			± 0.5	3.5	3.5	3.5			± 1.3	
	%	63.0	63.0	61.5			± 8.0	35.0	35.0	34.0			± 6.0	17.5	17.5	17.0			± 6.0	
MONO	#	0.2	0.2	0.2			± 0.2	0.6	0.6	0.6			± 0.4	1.0	1.0	1.0			± 1.0	
	%	12.0	12.0	11.5			± 7.0	7.5	7.5	7.5			± 5.0	5.0	5.0	5.0			± 5.0	
GRANULO	#	0.5	0.5	0.6			± 0.3	4.3	4.3	4.4			± 0.6	15.6	15.6	16.0			± 1.5	
	%	25.0	25.0	27.0			± 8.0	57.5	57.5	58.5			± 6.0	77.5	77.5	78.0			± 7.0	
CRP	mg/dl	0.76	0.70	N/A			± 0.30	2.47	2.34	N/A			± 0.70	4.18	4.00	N/A			± 1.00	
	mg/l	7.6	7.0	N/A			± 3.0	24.7	23.4	N/A			± 7.0	41.8	40.0	N/A			± 10.0	

Ref: TEMP-1070 Rev.36 FRONT/RECTO 9930089-A