

# ABX Minotrol CRP



**LOT** MC 013  
Rev 1

**CONTROL**

(Exp.) 2013-04-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse														
			CONTROL				1	CONTROL				2	CONTROL				3
			MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
			CRP	CRP 200	60 8P 60 18P			CRP	CRP 200	60 8P 60 18P			CRP	CRP 200	60 8P 60 18P		
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0	2.0		± 0.4	7.1	7.1	7.4		± 1.0	19.2	19.2	19.7		± 2.2
GR	RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.37	2.37	2.39		± 0.15	4.68	4.68	4.73		± 0.18	5.60	5.60	5.66		± 0.25
HB	HGB	g/dl	6.0	6.0	6.1		± 0.4	13.5	13.5	13.6		± 0.5	17.7	17.7	17.7		± 0.7
		g/l	60	60	61		± 4	135	135	136		± 5	177	177	177		± 7
		mmol/l	3.73	3.73	3.79		± 0.25	8.38	8.38	8.45		± 0.31	10.99	10.99	10.99		± 0.44
HT	HCT	%	16.6	16.6	16.7		± 2.0	38.4	38.4	38.8		± 2.5	49.8	49.8	50.4		± 3.0
		l/l	0.166	0.166	0.167		± 0.020	0.384	0.384	0.388		± 0.025	0.498	0.498	0.504		± 0.030
VGM	MCV	µm <sup>3</sup> ; fl	70	70	70		± 4	82	82	82		± 4	89	89	89		± 4
TGMH	MCH	pg	25.3	25.3	25.5		± 2.4	28.8	28.8	28.8		± 2.4	31.6	31.6	31.3		± 2.8
		fmol	1.57	1.57	1.58		± 0.15	1.79	1.79	1.79		± 0.15	1.96	1.96	1.94		± 0.18
CCMH	MCHC	g/dl	36.2	36.2	36.5		± 3.0	35.2	35.2	35.1		± 3.0	35.5	35.5	35.1		± 3.0
		g/l	362	362	365		± 30	352	352	351		± 30	355	355	351		± 30
		mmol/l	22.46	22.46	22.64		± 1.86	21.85	21.85	21.77		± 1.86	22.05	22.05	21.82		± 1.86
IDR	RDW	%	13.8	13.8	13.4		± 3.0	13.5	13.5	13.0		± 3.0	12.5	12.5	13.0		± 3.0
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	65	65	68		± 20	252	252	258		± 40	515	515	522		± 60
VPM	MPV	µm <sup>3</sup> ; fl	9.0	9.0	8.4		± 2.0	8.0	8.0	7.9		± 2.0	7.8	7.8	7.6		± 2.0
LYMPHO		#	1.3	1.3	1.3		± 0.2	2.3	2.3	2.4		± 0.5	2.4	2.4	2.6		± 1.1
		%	66.0	66.0	64.0		± 7.0	32.5	32.5	32.4		± 6.0	12.5	12.5	13.0		± 5.0
MONO		#	0.2	0.2	0.2		± 0.2	0.6	0.6	0.6		± 0.4	1.1	1.1	1.0		± 1.0
		%	9.0	9.0	9.0		± 7.0	8.0	8.0	7.8		± 5.0	5.5	5.5	5.0		± 5.0
GRANULO		#	0.5	0.5	0.5		± 0.2	4.2	4.2	4.4		± 0.5	15.7	15.7	16.2		± 1.4
		%	25.0	25.0	27.0		± 7.0	59.5	59.5	59.8		± 6.0	82.0	82.0	82.0		± 7.0
CRP		mg/dl	0.88	0.57	N/A		± 0.30	2.80	2.21	N/A		± 0.70	5.00	4.18	N/A		± 1.00
		mg/l	8.8	5.7	N/A		± 3.0	28.0	22.1	N/A		± 7.0	50.0	41.8	N/A		± 10.0

9930045-D

Ref. TEMP-1070 Rev.15