

ABX Minotrol CRP



LOT MC 101
Rev 2

CONTROL

(Exp.) 2012-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse												TOLERANCES TOLERANCE
			CONTROL			1	CONTROL			2	CONTROL			3	
			MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	
			CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0	2.0	± 0.4	7.4	7.4	7.5	± 1.0	19.8	19.8	20.1	± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.30	2.30	2.28	± 0.15	4.67	4.67	4.69	± 0.18	5.59	5.59	5.68	± 0.25		
HB HGB	g/dl	5.9	5.9	5.9	± 0.4	13.4	13.4	13.4	± 0.5	17.7	17.7	18.1	± 0.7		
	g/l	59	59	59	± 4	134	134	134	± 5	177	177	181	± 7		
HT HCT	mmol/l	3.66	3.66	3.66	± 0.25	8.32	8.32	8.32	± 0.31	10.99	10.99	11.24	± 0.44		
	%	16.6	16.6	16.4	± 2.0	37.8	37.8	38.5	± 2.5	50.9	50.9	52.3	± 3.0		
VGM MCV	l/l	0.166	0.166	0.164	± 0.020	0.378	0.378	0.385	± 0.025	0.509	0.509	0.523	± 0.030		
	µm ³ ; fl	72	72	72	± 4	81	81	82	± 4	91	91	92	± 4		
TGMH MCH	pg	25.7	25.7	25.9	± 2.4	28.7	28.7	28.6	± 2.4	31.7	31.7	31.9	± 2.8		
	fmol	1.59	1.59	1.61	± 0.15	1.78	1.78	1.77	± 0.15	1.97	1.97	1.98	± 0.18		
CCMH MCHC	g/dl	35.6	35.6	35.9	± 3.0	35.4	35.4	34.8	± 3.0	34.8	34.8	34.6	± 3.0		
	g/l	356	356	359	± 30	354	354	348	± 30	348	348	346	± 30		
	mmol/l	22.13	22.13	22.32	± 1.86	22.00	22.00	21.64	± 1.86	21.61	21.61	21.51	± 1.86		
IDR RDW	%	13.5	13.5	14.0	± 3.0	13.2	13.2	14.0	± 3.0	13.5	13.5	13.8	± 3.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	73	73	75	± 20	245	245	248	± 40	500	500	510	± 60		
VPM MPV	µm ³ ; fl	8.0	8.0	8.0	± 2.0	7.3	7.3	7.5	± 2.0	7.3	7.3	7.5	± 2.0		
LYMPHO	#	1.3	1.3	1.3	± 0.3	2.5	2.5	2.6	± 0.6	2.8	2.8	3.0	± 1.2		
	%	65.0	65.0	66.0	± 8.0	33.5	33.5	34.5	± 6.0	14.0	14.0	15.0	± 5.0		
MONO	#	0.2	0.2	0.2	± 0.2	0.6	0.6	0.5	± 0.5	1.2	1.2	1.0	± 1.0		
	%	11.0	11.0	10.0	± 7.0	7.5	7.5	6.5	± 5.0	6.0	6.0	5.0	± 5.0		
GRANULO	#	0.5	0.5	0.5	± 0.2	4.4	4.4	4.4	± 0.6	15.8	15.8	16.1	± 1.5		
	%	24.0	24.0	24.0	± 7.0	59.0	59.0	59.0	± 6.0	80.0	80.0	80.0	± 7.0		
CRP	mg/dl	0.64	0.48	N/A	± 0.30	2.60	2.35	N/A	± 0.70	4.20	4.10	N/A	± 1.00		
	mg/l	6.4	4.8	N/A	± 3.0	26.0	23.5	N/A	± 7.0	42.0	41.0	N/A	± 10.0		

Ref: FORM-1070 Rev.14 FRONT / RECTO 9930045-C