

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 043**

**CONTROL**



(Exp.) **2013-07-05**

(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse													TOLERANCES TOLERANCE
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P			
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>	<b>2.0</b>	± 0.4	<b>7.6</b>	<b>7.6</b>	<b>7.7</b>	± 1.0	<b>20.5</b>	<b>20.5</b>	<b>20.6</b>	± 2.2			
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.35</b>	<b>2.35</b>	<b>2.33</b>	± 0.15	<b>4.60</b>	<b>4.60</b>	<b>4.58</b>	± 0.18	<b>5.72</b>	<b>5.72</b>	<b>5.72</b>	± 0.25			
HB	HGB	g/dl	<b>6.1</b>	<b>6.1</b>	<b>6.1</b>	± 0.4	<b>13.5</b>	<b>13.5</b>	<b>13.5</b>	± 0.5	<b>18.0</b>	<b>18.0</b>	<b>18.0</b>	± 0.7			
		g/l	<b>61</b>	<b>61</b>	<b>61</b>	± 4	<b>135</b>	<b>135</b>	<b>135</b>	± 5	<b>180</b>	<b>180</b>	<b>180</b>	± 7			
		mmol/l	<b>3.79</b>	<b>3.79</b>	<b>3.79</b>	± 0.25	<b>8.38</b>	<b>8.38</b>	<b>8.38</b>	± 0.31	<b>11.18</b>	<b>11.18</b>	<b>11.18</b>	± 0.44			
HT	HCT	%	<b>16.7</b>	<b>16.7</b>	<b>16.5</b>	± 2.0	<b>37.7</b>	<b>37.7</b>	<b>37.6</b>	± 2.5	<b>50.9</b>	<b>50.9</b>	<b>50.9</b>	± 3.0			
		l/l	<b>0.167</b>	<b>0.167</b>	<b>0.165</b>	± 0.020	<b>0.377</b>	<b>0.377</b>	<b>0.376</b>	± 0.025	<b>0.509</b>	<b>0.509</b>	<b>0.509</b>	± 0.030			
VGM	MCV	µm <sup>3</sup> , fl	<b>71</b>	<b>71</b>	<b>71</b>	± 4	<b>82</b>	<b>82</b>	<b>82</b>	± 4	<b>89</b>	<b>89</b>	<b>89</b>	± 4			
TGMH	MCH	pg	<b>26.0</b>	<b>26.0</b>	<b>26.2</b>	± 2.4	<b>29.3</b>	<b>29.3</b>	<b>29.5</b>	± 2.4	<b>31.5</b>	<b>31.5</b>	<b>31.5</b>	± 2.8			
		fmol	<b>1.61</b>	<b>1.61</b>	<b>1.63</b>	± 0.15	<b>1.82</b>	<b>1.82</b>	<b>1.83</b>	± 0.15	<b>1.95</b>	<b>1.95</b>	<b>1.95</b>	± 0.18			
CCMH	MCHC	g/dl	<b>36.6</b>	<b>36.6</b>	<b>36.9</b>	± 3.0	<b>35.8</b>	<b>35.8</b>	<b>35.9</b>	± 3.0	<b>35.4</b>	<b>35.4</b>	<b>35.4</b>	± 3.0			
		g/l	<b>366</b>	<b>366</b>	<b>369</b>	± 30	<b>358</b>	<b>358</b>	<b>359</b>	± 30	<b>354</b>	<b>354</b>	<b>354</b>	± 30			
		mmol/l	<b>22.70</b>	<b>22.70</b>	<b>22.90</b>	± 1.86	<b>22.23</b>	<b>22.23</b>	<b>22.32</b>	± 1.86	<b>21.96</b>	<b>21.96</b>	<b>21.96</b>	± 1.86			
IDR	RDW	%	<b>13.6</b>	<b>13.6</b>	<b>13.8</b>	± 3.0	<b>13.8</b>	<b>13.8</b>	<b>13.2</b>	± 3.0	<b>13.0</b>	<b>13.0</b>	<b>13.0</b>	± 3.0			
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>73</b>	<b>73</b>	<b>74</b>	± 20	<b>260</b>	<b>260</b>	<b>260</b>	± 40	<b>515</b>	<b>515</b>	<b>510</b>	± 60			
VPM	MPV	µm <sup>3</sup> , fl	<b>8.5</b>	<b>8.5</b>	<b>8.3</b>	± 2.0	<b>7.8</b>	<b>7.8</b>	<b>7.5</b>	± 2.0	<b>7.8</b>	<b>7.8</b>	<b>7.6</b>	± 2.0			
LYMPHO		#	<b>1.3</b>	<b>1.3</b>	<b>1.3</b>	± 0.3	<b>2.5</b>	<b>2.5</b>	<b>2.5</b>	± 0.5	<b>3.0</b>	<b>3.0</b>	<b>3.0</b>	± 1.1			
		%	<b>64.0</b>	<b>64.0</b>	<b>62.5</b>	± 7.0	<b>32.3</b>	<b>32.3</b>	<b>32.0</b>	± 6.0	<b>14.5</b>	<b>14.5</b>	<b>14.5</b>	± 5.0			
MONO		#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>	± 0.2	<b>0.6</b>	<b>0.6</b>	<b>0.6</b>	± 0.4	<b>1.1</b>	<b>1.1</b>	<b>1.0</b>	± 1.0			
		%	<b>9.5</b>	<b>9.5</b>	<b>9.5</b>	± 7.0	<b>8.4</b>	<b>8.4</b>	<b>8.0</b>	± 5.0	<b>5.5</b>	<b>5.5</b>	<b>5.0</b>	± 5.0			
GRANULO		#	<b>0.5</b>	<b>0.5</b>	<b>0.6</b>	± 0.3	<b>4.5</b>	<b>4.5</b>	<b>4.6</b>	± 0.5	<b>16.4</b>	<b>16.4</b>	<b>16.6</b>	± 1.5			
		%	<b>26.5</b>	<b>26.5</b>	<b>28.0</b>	± 7.0	<b>59.3</b>	<b>59.3</b>	<b>60.0</b>	± 6.0	<b>80.0</b>	<b>80.0</b>	<b>80.5</b>	± 7.0			
CRP		mg/dl	<b>1.00</b>	<b>0.60</b>	N/A	± 0.30	<b>2.80</b>	<b>2.10</b>	N/A	± 0.70	<b>4.80</b>	<b>4.00</b>	N/A	± 1.00			
		mg/l	<b>10.0</b>	<b>6.0</b>	N/A	± 3.0	<b>28.0</b>	<b>21.0</b>	N/A	± 7.0	<b>48.0</b>	<b>40.0</b>	N/A	± 10.0			

9930074-A FRONT/RECTO Ref: TEMP-1070 Rev.31