

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 016**

**CONTROL**

(Exp.) **2016-04-05**  
(YYYY - MM - DD)

Rev 3

		<b>ABX Lyse</b>																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>	<b>2.0</b>		± 0.4	<b>7.2</b>	<b>7.2</b>	<b>7.3</b>		± 1.0	<b>19.3</b>	<b>19.3</b>	<b>19.7</b>		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.45</b>	<b>2.45</b>	<b>2.44</b>		± 0.15	<b>4.65</b>	<b>4.65</b>	<b>4.65</b>		± 0.18	<b>5.62</b>	<b>5.62</b>	<b>5.62</b>		± 0.25			
HB HGB	g/dl	<b>6.1</b>	<b>6.1</b>	<b>6.1</b>		± 0.4	<b>13.4</b>	<b>13.4</b>	<b>13.5</b>		± 0.5	<b>18.0</b>	<b>18.0</b>	<b>18.0</b>		± 0.7			
	g/l	<b>61</b>	<b>61</b>	<b>61</b>		± 4	<b>134</b>	<b>134</b>	<b>135</b>		± 5	<b>180</b>	<b>180</b>	<b>180</b>		± 7			
	mmol/l	<b>3.79</b>	<b>3.79</b>	<b>3.79</b>		± 0.25	<b>8.32</b>	<b>8.32</b>	<b>8.38</b>		± 0.31	<b>11.18</b>	<b>11.18</b>	<b>11.18</b>		± 0.44			
HT HCT	%	<b>16.9</b>	<b>16.9</b>	<b>16.6</b>		± 2.0	<b>37.2</b>	<b>37.2</b>	<b>37.2</b>		± 2.5	<b>50.0</b>	<b>50.0</b>	<b>50.0</b>		± 3.0			
	l/l	<b>0.169</b>	<b>0.169</b>	<b>0.166</b>		± 0.020	<b>0.372</b>	<b>0.372</b>	<b>0.372</b>		± 0.025	<b>0.500</b>	<b>0.500</b>	<b>0.500</b>		± 0.030			
VGM MCV	µm <sup>3</sup> ; fl	<b>69</b>	<b>69</b>	<b>68</b>		± 4	<b>80</b>	<b>80</b>	<b>80</b>		± 4	<b>89</b>	<b>89</b>	<b>89</b>		± 4			
TGMH MCH	pg	<b>24.9</b>	<b>24.9</b>	<b>25.0</b>		± 2.4	<b>28.8</b>	<b>28.8</b>	<b>29.0</b>		± 2.4	<b>32.0</b>	<b>32.0</b>	<b>32.0</b>		± 2.8			
	fmol	<b>1.55</b>	<b>1.55</b>	<b>1.55</b>		± 0.15	<b>1.79</b>	<b>1.79</b>	<b>1.80</b>		± 0.15	<b>1.99</b>	<b>1.99</b>	<b>1.99</b>		± 0.18			
CCMH MCHC	g/dl	<b>36.1</b>	<b>36.1</b>	<b>36.8</b>		± 3.0	<b>36.0</b>	<b>36.0</b>	<b>36.3</b>		± 3.0	<b>36.0</b>	<b>36.0</b>	<b>36.0</b>		± 3.0			
	g/l	<b>361</b>	<b>361</b>	<b>368</b>		± 30	<b>360</b>	<b>360</b>	<b>363</b>		± 30	<b>360</b>	<b>360</b>	<b>360</b>		± 30			
	mmol/l	<b>22.41</b>	<b>22.41</b>	<b>22.83</b>		± 1.86	<b>22.37</b>	<b>22.37</b>	<b>22.54</b>		± 1.86	<b>22.35</b>	<b>22.35</b>	<b>22.35</b>		± 1.86			
IDR RDW	%	<b>13.5</b>	<b>13.5</b>	<b>14.0</b>		± 3.0	<b>13.0</b>	<b>13.0</b>	<b>13.5</b>		± 3.0	<b>12.5</b>	<b>12.5</b>	<b>13.0</b>		± 3.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>75</b>	<b>75</b>	<b>75</b>		± 20	<b>250</b>	<b>250</b>	<b>252</b>		± 40	<b>495</b>	<b>495</b>	<b>496</b>		± 60			
VPM MPV	µm <sup>3</sup> ; fl	<b>9.0</b>	<b>9.0</b>	<b>8.5</b>		± 2.0	<b>8.0</b>	<b>8.0</b>	<b>7.5</b>		± 2.0	<b>7.8</b>	<b>7.8</b>	<b>7.5</b>		± 2.0			
LYMPHO	#	<b>1.3</b>	<b>1.3</b>	<b>1.3</b>		± 0.4	<b>2.4</b>	<b>2.4</b>	<b>2.5</b>		± 0.5	<b>2.9</b>	<b>2.9</b>	<b>3.1</b>		± 1.2			
	%	<b>67.0</b>	<b>67.0</b>	<b>65.0</b>		± 8.0	<b>34.0</b>	<b>34.0</b>	<b>33.5</b>		± 6.0	<b>15.0</b>	<b>15.0</b>	<b>15.5</b>		± 6.0			
MONO	#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>		± 0.2	<b>0.6</b>	<b>0.6</b>	<b>0.6</b>		± 0.4	<b>1.2</b>	<b>1.2</b>	<b>1.1</b>		± 1.0			
	%	<b>10.5</b>	<b>10.5</b>	<b>11.0</b>		± 7.0	<b>8.5</b>	<b>8.5</b>	<b>8.5</b>		± 5.0	<b>6.0</b>	<b>6.0</b>	<b>5.5</b>		± 5.0			
GRANULO	#	<b>0.5</b>	<b>0.5</b>	<b>0.5</b>		± 0.4	<b>4.2</b>	<b>4.2</b>	<b>4.2</b>		± 0.5	<b>15.2</b>	<b>15.2</b>	<b>15.5</b>		± 1.4			
	%	<b>22.5</b>	<b>22.5</b>	<b>24.0</b>		± 8.0	<b>57.5</b>	<b>57.5</b>	<b>58.0</b>		± 6.0	<b>79.0</b>	<b>79.0</b>	<b>79.0</b>		± 7.0			
CRP	mg/dl	<b>0.71</b>	<b>0.60</b>	N/A		± 0.30	<b>2.63</b>	<b>2.53</b>	N/A		± 0.70	<b>4.32</b>	<b>4.20</b>	N/A		± 1.00			
	mg/l	<b>7.1</b>	<b>6.0</b>	N/A		± 3.0	<b>26.3</b>	<b>25.3</b>	N/A		± 7.0	<b>43.2</b>	<b>42.0</b>	N/A		± 10.0			

Ref: TEMP-1070 Rev.36 FRONT/RECTO 9930089-A