

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 073**

**CONTROL**



**(Exp.) 2013-10-05**

(YYYY - MM - DD)

Rev 1

		<b>ABX Lyse</b>																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					TOLERANCES TOLERANCE	CONTROL					TOLERANCES TOLERANCE	CONTROL					TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>1.9</b>	<b>1.9</b>	<b>2.0</b>			± 0.4	<b>7.3</b>	<b>7.3</b>	<b>7.6</b>			± 1.0	<b>19.5</b>	<b>19.5</b>	<b>19.7</b>			± 2.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.35</b>	<b>2.35</b>	<b>2.36</b>			± 0.15	<b>4.58</b>	<b>4.58</b>	<b>4.58</b>			± 0.18	<b>5.51</b>	<b>5.51</b>	<b>5.54</b>			± 0.25
HB HGB	g/dl	<b>6.0</b>	<b>6.0</b>	<b>6.0</b>			± 0.4	<b>13.2</b>	<b>13.2</b>	<b>13.2</b>			± 0.5	<b>17.7</b>	<b>17.7</b>	<b>17.8</b>			± 0.7
	g/l	<b>60</b>	<b>60</b>	<b>60</b>			± 4	<b>132</b>	<b>132</b>	<b>132</b>			± 5	<b>177</b>	<b>177</b>	<b>178</b>			± 7
	mmol/l	<b>3.73</b>	<b>3.73</b>	<b>3.73</b>			± 0.25	<b>8.20</b>	<b>8.20</b>	<b>8.20</b>			± 0.31	<b>10.99</b>	<b>10.99</b>	<b>11.05</b>			± 0.44
HT HCT	%	<b>16.5</b>	<b>16.5</b>	<b>16.5</b>			± 2.0	<b>37.1</b>	<b>37.1</b>	<b>37.1</b>			± 2.5	<b>49.6</b>	<b>49.6</b>	<b>49.9</b>			± 3.0
	l/l	<b>0.165</b>	<b>0.165</b>	<b>0.165</b>			± 0.020	<b>0.371</b>	<b>0.371</b>	<b>0.371</b>			± 0.025	<b>0.496</b>	<b>0.496</b>	<b>0.499</b>			± 0.030
VGM MCV	µm <sup>3</sup> , fl	<b>70</b>	<b>70</b>	<b>70</b>			± 4	<b>81</b>	<b>81</b>	<b>81</b>			± 4	<b>90</b>	<b>90</b>	<b>90</b>			± 4
TGMH MCH	pg	<b>25.5</b>	<b>25.5</b>	<b>25.4</b>			± 2.4	<b>28.8</b>	<b>28.8</b>	<b>28.8</b>			± 2.4	<b>32.1</b>	<b>32.1</b>	<b>32.1</b>			± 2.8
	fmol	<b>1.59</b>	<b>1.59</b>	<b>1.58</b>			± 0.15	<b>1.79</b>	<b>1.79</b>	<b>1.79</b>			± 0.15	<b>1.99</b>	<b>1.99</b>	<b>2.00</b>			± 0.18
CCMH MCHC	g/dl	<b>36.5</b>	<b>36.5</b>	<b>36.3</b>			± 3.0	<b>35.6</b>	<b>35.6</b>	<b>35.6</b>			± 3.0	<b>35.7</b>	<b>35.7</b>	<b>35.7</b>			± 3.0
	g/l	<b>365</b>	<b>365</b>	<b>363</b>			± 30	<b>356</b>	<b>356</b>	<b>356</b>			± 30	<b>357</b>	<b>357</b>	<b>357</b>			± 30
	mmol/l	<b>22.65</b>	<b>22.65</b>	<b>22.55</b>			± 1.86	<b>22.10</b>	<b>22.10</b>	<b>22.10</b>			± 1.86	<b>22.17</b>	<b>22.17</b>	<b>22.17</b>			± 1.86
IDR RDW	%	<b>13.8</b>	<b>13.8</b>	<b>13.9</b>			± 3.0	<b>13.6</b>	<b>13.6</b>	<b>13.4</b>			± 3.0	<b>12.7</b>	<b>12.7</b>	<b>12.8</b>			± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>75</b>	<b>75</b>	<b>75</b>			± 20	<b>250</b>	<b>250</b>	<b>250</b>			± 40	<b>500</b>	<b>500</b>	<b>500</b>			± 60
VPM MPV	µm <sup>3</sup> , fl	<b>9.1</b>	<b>9.1</b>	<b>8.9</b>			± 2.0	<b>8.1</b>	<b>8.1</b>	<b>7.8</b>			± 2.0	<b>8.1</b>	<b>8.1</b>	<b>7.9</b>			± 2.0
LYMPHO	#	<b>1.2</b>	<b>1.2</b>	<b>1.3</b>			± 0.3	<b>2.4</b>	<b>2.4</b>	<b>2.6</b>			± 0.5	<b>2.8</b>	<b>2.8</b>	<b>3.0</b>			± 1.2
	%	<b>64.0</b>	<b>64.0</b>	<b>62.5</b>			± 8.0	<b>33.5</b>	<b>33.5</b>	<b>34.5</b>			± 6.0	<b>14.5</b>	<b>14.5</b>	<b>15.0</b>			± 6.0
MONO	#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>			± 0.2	<b>0.5</b>	<b>0.5</b>	<b>0.5</b>			± 0.4	<b>1.1</b>	<b>1.1</b>	<b>1.0</b>			± 1.0
	%	<b>10.0</b>	<b>10.0</b>	<b>10.5</b>			± 7.0	<b>7.5</b>	<b>7.5</b>	<b>6.5</b>			± 5.0	<b>5.5</b>	<b>5.5</b>	<b>5.0</b>			± 5.0
GRANULO	#	<b>0.5</b>	<b>0.5</b>	<b>0.5</b>			± 0.2	<b>4.3</b>	<b>4.3</b>	<b>4.5</b>			± 0.6	<b>15.6</b>	<b>15.6</b>	<b>15.8</b>			± 1.4
	%	<b>26.0</b>	<b>26.0</b>	<b>27.0</b>			± 7.0	<b>59.0</b>	<b>59.0</b>	<b>59.0</b>			± 6.0	<b>80.0</b>	<b>80.0</b>	<b>80.0</b>			± 7.0
CRP	mg/dl	<b>1.08</b>	<b>0.69</b>	N/A			± 0.30	<b>2.85</b>	<b>2.24</b>	N/A			± 0.70	<b>4.77</b>	<b>4.03</b>	N/A			± 1.00
	mg/l	<b>10.8</b>	<b>6.9</b>	N/A			± 3.0	<b>28.5</b>	<b>22.4</b>	N/A			± 7.0	<b>47.7</b>	<b>40.3</b>	N/A			± 10.0

9930074-A  
FRONT/RECTO  
Rev.31  
Ref: TEMP-1070