

ABX Minotrol CRP



LOT

MC 075

Rev 2

CONTROL



(Exp.) 2015-10-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse																
				CONTROL				1	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
				MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE		MICROS				TOLERANCES TOLERANCE	
				CRP	CRP 200	60 8P 60 18P			CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P			
GB	WBC	10 ³ /mm ³ ; 10 ⁹ /l	2,0	2,0	2,0			± 0,4	7,6	7,6	7,6			± 1,0	20,2	20,2	20,1			± 2,2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2,41	2,41	2,42			± 0,15	4,60	4,60	4,60			± 0,18	5,68	5,68	5,75			± 0,25
HB	HGB	g/dl	6,1	6,1	6,1			± 0,4	13,5	13,5	13,4			± 0,5	17,8	17,8	17,9			± 0,7
		g/l	61	61	61			± 4	135	135	134			± 5	178	178	179			± 7
HT	HCT	mmol/l	3,79	3,79	3,79			± 0,25	8,38	8,38	8,32			± 0,31	11,05	11,05	11,12			± 0,44
		%	16,6	16,6	16,7			± 2,0	37,3	37,3	37,7			± 2,5	50,0	50,0	50,6			± 3,0
VGM	MCV	l/l	0,166	0,166	0,167			± 0,020	0,373	0,373	0,377			± 0,025	0,500	0,500	0,506			± 0,030
		µm ³ ; fl	69	69	69			± 4	81	81	82			± 4	88	88	88			± 4
TGMH	MCH	pg	25,3	25,3	25,2			± 2,4	29,3	29,3	29,1			± 2,4	31,3	31,3	31,1			± 2,8
		fmol	1,57	1,57	1,57			± 0,15	1,82	1,82	1,81			± 0,15	1,95	1,95	1,93			± 0,18
CCMH	MCHC	g/dl	36,7	36,7	36,5			± 3,0	36,2	36,2	35,5			± 3,0	35,6	35,6	35,4			± 3,0
		g/l	367	367	365			± 30	362	362	355			± 30	356	356	354			± 30
IDR	RDW	mmol/l	22,78	22,78	22,69			± 1,86	22,50	22,50	22,06			± 1,86	22,11	22,11	21,97			± 1,86
		%	13,8	13,8	14,0			± 3,0	13,5	13,5	13,6			± 3,0	13,2	13,2	13,0			± 3,0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	72	75			± 20	246	246	250			± 40	510	510	520			± 60
VPM	MPV	µm ³ ; fl	9,0	9,0	8,3			± 2,0	8,3	8,3	7,9			± 2,0	8,0	8,0	7,7			± 2,0
LYMPHO		#	1,3	1,3	1,3			± 0,3	2,6	2,6	2,5			± 0,5	3,0	3,0	2,9			± 1,3
		%	64,5	64,5	63,0			± 8,0	34,0	34,0	33,5			± 6,0	15,0	15,0	14,5			± 6,0
MONO		#	0,2	0,2	0,2			± 0,2	0,6	0,6	0,6			± 0,5	1,2	1,2	1,1			± 1,0
		%	11,5	11,5	11,0			± 7,0	8,5	8,5	8,0			± 5,0	6,0	6,0	5,5			± 5,0
GRANULO		#	0,5	0,5	0,5			± 0,2	4,4	4,4	4,4			± 0,6	16,0	16,0	16,1			± 1,5
		%	24,0	24,0	26,0			± 7,0	57,5	57,5	58,5			± 6,0	79,0	79,0	80,0			± 7,0
CRP		mg/dl	0,88	0,59	N/A			± 0,30	2,56	2,09	N/A			± 0,70	4,79	4,23	N/A			± 1,00
		mg/l	8,8	5,9	N/A			± 3,0	25,6	20,9	N/A			± 7,0	47,9	42,3	N/A			± 10,0

Ref: TEMP-1070 Rev.36 FRONT/RECTO 9930089-A