

ABX Minotrol CRP



LOT

MC 412

CONTROL



(Exp.) **2019-10-05**
(YYYY-MM-DD)

Rev 1

		Minilysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	1.9	2.0			± 0.4	7.4	7.7			± 1.0	19.3	21.0			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.31	2.31			± 0.15	4.59	4.59			± 0.18	5.55	5.58			± 0.25	
HB HGB	g/dl	5.7	5.8			± 0.4	13.0	13.2			± 0.5	17.6	18.2			± 0.7	
	g/l	57	58			± 4	130	132			± 5	176	182			± 7	
	mmol/l	3.54	3.60			± 0.25	8.09	8.20			± 0.31	10.93	11.30			± 0.44	
HT HCT	%	16.2	16.2			± 2.0	36.7	36.7			± 2.5	49.4	50.2			± 3.0	
	l/l	0.162	0.162			± 0.020	0.367	0.367			± 0.025	0.494	0.502			± 0.030	
VGM MCV	µm ³ ; fl	70	70			± 4	80	80			± 4	89	90			± 4	
TGMH MCH	pg	24.7	25.1			± 2.4	28.4	28.8			± 2.4	31.7	32.6			± 2.8	
	fmol	1.53	1.56			± 0.15	1.76	1.79			± 0.15	1.97	2.03			± 0.18	
CCMH MCHC	g/dl	35.3	35.9			± 3.0	35.5	35.9			± 3.0	35.6	36.2			± 3.0	
	g/l	353	359			± 30	355	359			± 30	356	362			± 30	
	mmol/l	21.89	22.27			± 1.86	22.04	22.32			± 1.86	22.13	22.51			± 1.86	
IDR RDW	%	14.4	14.0			± 3.0	13.8	14.0			± 3.0	12.8	13.5			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	76	75			± 20	258	260			± 40	510	520			± 60	
VPM MPV	µm ³ ; fl	8.1	8.9			± 2.0	7.9	8.5			± 2.0	7.7	8.4			± 2.0	
LYMPHO	#	1.3	1.3			± 0.2	2.9	2.8			± 0.5	3.6	3.8			± 1.3	
	%	68.6	65.5			± 7.0	39.4	36.5			± 6.0	18.7	18.0			± 6.0	
MONO	#	0.1	0.2			± 0.2	0.5	0.5			± 0.4	1.2	1.1			± 0.8	
	%	7.6	7.5			± 6.0	7.1	6.0			± 5.0	6.0	5.0			± 4.0	
GRANULO	#	0.5	0.5			± 0.2	4.0	4.4			± 0.5	14.6	16.2			± 1.5	
	%	24.0	27.0			± 7.0	53.9	57.5			± 6.0	75.7	77.0			± 7.0	
CRP	mg/dl	0.48	N/A			± 0.30	2.10	N/A			± 0.70	4.12	N/A			± 1.00	
	mg/l	4.8	N/A			± 3.0	21.0	N/A			± 7.0	41.2	N/A			± 10.0	

Ref: TEMP-1070 Rev.38 1300027267-A