

ABX Minotrol CRP



LOT

MC 419

CONTROL



(Exp.) **2021-07-05**
(YYYY - MM - DD)

Rev 1

		Minilysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.1			± 0.4	7.5	7.6			± 1.0	19.8	20.7			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.32	2.38			± 0.15	4.51	4.54			± 0.18	5.51	5.55			± 0.25	
HB HGB	g/dl	5.6	5.9			± 0.4	12.8	13.3			± 0.5	17.8	18.3			± 0.7	
	g/l	56	59			± 4	128	133			± 5	178	183			± 7	
	mmol/l	3.48	3.66			± 0.25	7.97	8.26			± 0.31	11.05	11.36			± 0.44	
HT HCT	%	16.2	16.9			± 2.0	36.5	37.2			± 2.5	50.1	51.1			± 3.0	
	l/l	0.162	0.169			± 0.020	0.365	0.372			± 0.025	0.501	0.511			± 0.030	
VGM MCV	µm ³ ; fl	70	71			± 4	81	82			± 4	91	92			± 4	
TGMH MCH	pg	24.2	24.8			± 2.4	28.5	29.3			± 2.4	32.3	33.0			± 2.8	
	fmol	1.50	1.54			± 0.15	1.77	1.82			± 0.15	2.01	2.05			± 0.18	
CCMH MCHC	g/dl	34.5	34.9			± 3.0	35.1	35.7			± 3.0	35.5	35.8			± 3.0	
	g/l	345	349			± 30	351	357			± 30	355	358			± 30	
	mmol/l	21.43	21.68			± 1.86	21.82	22.19			± 1.86	22.05	22.26			± 1.86	
IDR RDW	%	13.6	14.5			± 3.0	13.8	14.5			± 3.0	13.0	14.0			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	67	75			± 20	254	258			± 40	510	515			± 60	
VPM MPV	µm ³ ; fl	8.1	8.6			± 2.0	7.4	8.4			± 2.0	7.2	8.0			± 2.0	
LYMPHO	#	1.4	1.4			± 0.2	2.8	2.7			± 0.5	3.9	3.5			± 1.1	
	%	72.3	68.0			± 8.0	37.3	35.0			± 6.0	19.8	17.0			± 5.0	
MONO	#	0.2	0.2			± 0.2	0.5	0.5			± 0.4	1.0	1.0			± 1.0	
	%	5.6	6.5			± 5.6	7.1	6.5			± 5.0	5.0	5.0			± 5.0	
GRANULO	#	0.5	0.5			± 0.2	4.2	4.4			± 0.5	15.0	16.1			± 1.5	
	%	23.0	25.5			± 7.0	55.9	58.5			± 6.0	75.7	78.0			± 7.0	
CRP	mg/dl	0.55	N/A			± 0.30	2.20	N/A			± 0.70	4.10	N/A			± 1.00	
	mg/l	5.5	N/A			± 3.0	22.0	N/A			± 7.0	41.0	N/A			± 10.0	

1300027267-A
Ref: TEMP-1070 Rev.38