

ABX Minotrol CRP



LOT

MC 417

CONTROL



(Exp.) **2021-01-05**
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE		
				CONTROL			1	CONTROL			2		CONTROL					3
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE			
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P					
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0		± 0.4	7.6	7.7		± 1.0	19.4	20.1		± 2.2				
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.42	2.37		± 0.15	4.56	4.52		± 0.18	5.50	5.48		± 0.25				
HB	HGB	g/dl	5.8	5.8		± 0.4	13.2	13.3		± 0.5	17.7	18.0		± 0.7				
		g/l	58	58		± 4	132	133		± 5	177	180		± 7				
		mmol/l	3.60	3.60		± 0.25	8.22	8.26		± 0.31	10.99	11.18		± 0.44				
HT	HCT	%	16.9	17.1		± 2.0	37.8	37.5		± 2.5	50.1	50.4		± 3.0				
		l/l	0.169	0.171		± 0.020	0.378	0.375		± 0.025	0.501	0.504		± 0.030				
VGM	MCV	µm ³ ; fl	70	72		± 4	83	83		± 4	91	92		± 4				
TGMH	MCH	pg	23.9	24.5		± 2.4	29.0	29.4		± 2.4	32.2	32.8		± 2.8				
		fmol	1.49	1.52		± 0.15	1.80	1.83		± 0.15	2.00	2.04		± 0.18				
CCMH	MCHC	g/dl	34.2	34.0		± 3.0	35.0	35.5		± 3.0	35.4	35.7		± 3.0				
		g/l	342	340		± 30	350	355		± 30	354	357		± 30				
		mmol/l	21.24	21.11		± 1.86	21.71	22.02		± 1.86	21.96	22.17		± 1.86				
IDR	RDW	%	14.0	14.5		± 3.0	13.5	14.2		± 3.0	12.5	13.5		± 3.0				
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	74	75		± 20	252	255		± 40	507	510		± 60				
VPM	MPV	µm ³ ; fl	8.9	8.9		± 2.0	7.9	8.4		± 2.0	7.3	7.8		± 2.0				
LYMPHO	#		1.4	1.3		± 0.2	2.8	2.8		± 0.5	3.5	3.5		± 1.2				
	%		68.6	67.0		± 8.0	37.3	36.5		± 6.0	18.2	17.5		± 6.0				
MONO	#		0.2	0.2		± 0.2	0.6	0.5		± 0.4	1.0	0.8		± 0.8				
	%		8.4	8.0		± 7.0	7.4	6.5		± 5.0	5.0	4.0		± 4.0				
GRANULO	#		0.5	0.5		± 0.2	4.2	4.4		± 0.5	15.0	15.8		± 1.5				
	%		23.0	25.0		± 7.0	55.6	57.0		± 6.0	77.2	78.5		± 7.0				
CRP	mg/dl		0.46	N/A		± 0.30	1.92	N/A		± 0.70	3.94	N/A		± 1.00				
	mg/l		4.6	N/A		± 3.0	19.2	N/A		± 7.0	39.4	N/A		± 10.0				

1300027267-A
Ref: TEMP-1070 Rev.38