

# ABX Minotrol CRP



**LOT**

**MC 407**

**CONTROL**



**(Exp.) 2018-07-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>		± 0.4	<b>7.5</b>	<b>7.7</b>		± 1.0	<b>19.7</b>	<b>20.4</b>		± 2.2				
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.32</b>	<b>2.31</b>		± 0.15	<b>4.62</b>	<b>4.62</b>		± 0.18	<b>5.83</b>	<b>5.86</b>		± 0.25				
HB HGB	g/dl	<b>5.8</b>	<b>5.9</b>		± 0.4	<b>13.0</b>	<b>13.4</b>		± 0.5	<b>17.7</b>	<b>18.2</b>		± 0.7				
	g/l	<b>58</b>	<b>59</b>		± 4	<b>130</b>	<b>134</b>		± 5	<b>177</b>	<b>182</b>		± 7				
HT HCT	mmol/l	<b>3.60</b>	<b>3.66</b>		± 0.25	<b>8.09</b>	<b>8.32</b>		± 0.31	<b>10.99</b>	<b>11.30</b>		± 0.44				
	%	<b>16.2</b>	<b>16.2</b>		± 2.0	<b>37.4</b>	<b>37.4</b>		± 2.5	<b>50.1</b>	<b>50.4</b>		± 3.0				
VGM MCV	l/l	<b>0.162</b>	<b>0.162</b>		± 0.020	<b>0.374</b>	<b>0.374</b>		± 0.025	<b>0.501</b>	<b>0.504</b>		± 0.030				
	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>		± 4	<b>81</b>	<b>81</b>		± 4	<b>86</b>	<b>86</b>		± 4				
TGMH MCH	pg	<b>25.0</b>	<b>25.5</b>		± 2.4	<b>28.2</b>	<b>29.0</b>		± 2.4	<b>30.4</b>	<b>31.1</b>		± 2.8				
	fmol	<b>1.55</b>	<b>1.59</b>		± 0.15	<b>1.75</b>	<b>1.80</b>		± 0.15	<b>1.89</b>	<b>1.93</b>		± 0.18				
CCMH MCHC	g/dl	<b>35.7</b>	<b>36.5</b>		± 3.0	<b>34.8</b>	<b>35.8</b>		± 3.0	<b>35.3</b>	<b>36.1</b>		± 3.0				
	g/l	<b>357</b>	<b>365</b>		± 30	<b>348</b>	<b>358</b>		± 30	<b>353</b>	<b>361</b>		± 30				
	mmol/l	<b>22.16</b>	<b>22.66</b>		± 1.86	<b>21.63</b>	<b>22.24</b>		± 1.86	<b>21.92</b>	<b>22.43</b>		± 1.86				
IDR RDW	%	<b>14.0</b>	<b>13.8</b>		± 3.0	<b>14.0</b>	<b>14.2</b>		± 3.0	<b>13.0</b>	<b>13.4</b>		± 3.0				
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>79</b>	<b>86</b>		± 20	<b>256</b>	<b>265</b>		± 40	<b>509</b>	<b>520</b>		± 60				
VPM MPV	µm <sup>3</sup> ; fl	<b>8.1</b>	<b>8.2</b>		± 2.0	<b>8.5</b>	<b>8.9</b>		± 2.0	<b>8.0</b>	<b>8.4</b>		± 2.0				
LYMPHO	#	<b>1.4</b>	<b>1.3</b>		± 0.2	<b>2.6</b>	<b>2.7</b>		± 0.5	<b>3.2</b>	<b>3.2</b>		± 1.2				
	%	<b>68.6</b>	<b>67.0</b>		± 8.0	<b>35.2</b>	<b>34.5</b>		± 6.0	<b>16.1</b>	<b>15.5</b>		± 6.0				
MONO	#	<b>0.2</b>	<b>0.2</b>		± 0.2	<b>0.5</b>	<b>0.4</b>		± 0.4	<b>0.9</b>	<b>0.8</b>		± 0.8				
	%	<b>8.8</b>	<b>9.0</b>		± 7.0	<b>6.7</b>	<b>5.5</b>		± 5.0	<b>4.5</b>	<b>4.0</b>		± 4.0				
GRANULO	#	<b>0.5</b>	<b>0.5</b>		± 0.2	<b>4.4</b>	<b>4.6</b>		± 0.5	<b>15.7</b>	<b>16.4</b>		± 1.5				
	%	<b>22.5</b>	<b>24.0</b>		± 7.0	<b>58.3</b>	<b>60.0</b>		± 6.0	<b>79.7</b>	<b>80.5</b>		± 7.0				
CRP	mg/dl	<b>0.50</b>	N/A		± 0.30	<b>2.65</b>	N/A		± 0.70	<b>3.88</b>	N/A		± 1.00				
	mg/l	<b>5.0</b>	N/A		± 3.0	<b>26.5</b>	N/A		± 7.0	<b>38.8</b>	N/A		± 10.0				

Ref: TEMP-1070 Rev.37 1300027267-A