

ABX Minotrol CRP



LOT

MC 416

CONTROL



(Exp.) 2020-10-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	1.9	2.1		± 0.4	7.6	7.8		± 1.0	19.6	20.5		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.32	2.35		± 0.15	4.50	4.54		± 0.18	5.58	5.62		± 0.25				
HB HGB	g/dl	5.7	5.8		± 0.4	12.9	13.2		± 0.5	17.7	18.2		± 0.7				
	g/l	57	58		± 4	129	132		± 5	177	182		± 7				
HT HCT	mmol/l	3.54	3.60		± 0.25	8.03	8.20		± 0.31	10.99	11.30		± 0.44				
	%	16.2	16.5		± 2.0	36.9	36.8		± 2.5	49.7	50.0		± 3.0				
VGM MCV	l/l	0.162	0.165		± 0.020	0.369	0.368		± 0.025	0.497	0.500		± 0.030				
	µm ³ ; fl	70	70		± 4	82	81		± 4	89	89		± 4				
TGMH MCH	pg	24.6	24.7		± 2.4	28.7	29.1		± 2.4	31.7	32.4		± 2.8				
	fmol	1.53	1.53		± 0.15	1.79	1.81		± 0.15	1.97	2.01		± 0.18				
CCMH MCHC	g/dl	35.1	35.3		± 3.0	35.1	35.9		± 3.0	35.6	36.4		± 3.0				
	g/l	351	353		± 30	351	359		± 30	356	364		± 30				
	mmol/l	21.80	21.90		± 1.86	21.77	22.29		± 1.86	22.13	22.60		± 1.86				
IDR RDW	%	14.0	14.0		± 3.0	13.5	14.2		± 3.0	13.0	13.8		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	68		± 20	265	265		± 40	505	515		± 60				
VPM MPV	µm ³ ; fl	8.0	8.9		± 2.0	7.5	8.2		± 2.0	7.3	8.0		± 2.0				
LYMPHO	#	1.3	1.3		± 0.2	2.7	2.7		± 0.5	3.6	3.5		± 1.3				
	%	66.0	63.5		± 8.0	35.7	34.0		± 6.0	18.2	17.0		± 6.0				
MONO	#	0.2	0.2		± 0.2	0.5	0.5		± 0.4	1.0	0.8		± 0.8				
	%	8.4	10.0		± 7.0	7.1	6.5		± 5.0	5.0	4.0		± 4.0				
GRANULO	#	0.5	0.6		± 0.2	4.4	4.6		± 0.5	15.1	16.2		± 1.5				
	%	25.5	26.5		± 7.0	57.3	59.5		± 6.0	77.2	79.0		± 7.0				
CRP	mg/dl	0.42	N/A		± 0.30	2.05	N/A		± 0.70	3.55	N/A		± 1.00				
	mg/l	4.2	N/A		± 3.0	20.5	N/A		± 7.0	35.5	N/A		± 10.0				

Ref: TEMP-1070 Rev.38 1300027267-A