

ABX Minotrol CRP



LOT

MC 415

CONTROL



(Exp.) **2020-07-05**

(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio											
				CONTROL			1	CONTROL			2	CONTROL			3
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P		
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.1		± 0.4	7.5	7.7		± 1.0	20.0	21.1		± 2.2		
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.43	2.42		± 0.15	4.61	4.67		± 0.18	5.62	5.72		± 0.25		
HB HGB	g/dl	5.9	6.0		± 0.4	13.0	13.5		± 0.5	18.0	18.8		± 0.7		
	g/l	59	60		± 4	130	135		± 5	180	188		± 7		
HT HCT	mmol/l	3.66	3.73		± 0.25	8.09	8.38		± 0.31	11.18	11.67		± 0.44		
	%	16.8	16.9		± 2.0	36.9	37.8		± 2.5	50.6	51.5		± 3.0		
VGM MCV	l/l	0.168	0.169		± 0.020	0.369	0.378		± 0.025	0.506	0.515		± 0.030		
	µm ³ ; fl	69	70		± 4	80	81		± 4	90	90		± 4		
TGMH MCH	pg	24.2	24.8		± 2.4	28.3	28.9		± 2.4	32.0	32.9		± 2.8		
	fmol	1.51	1.54		± 0.15	1.76	1.80		± 0.15	1.99	2.04		± 0.18		
CCMH MCHC	g/dl	35.1	35.4		± 3.0	35.3	35.7		± 3.0	35.6	36.5		± 3.0		
	g/l	351	354		± 30	353	357		± 30	356	365		± 30		
	mmol/l	21.81	22.00		± 1.86	21.95	22.16		± 1.86	22.10	22.68		± 1.86		
IDR RDW	%	13.7	14.6		± 3.0	14.6	14.7		± 3.0	13.5	14.4		± 3.0		
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	65	70		± 20	245	250		± 40	500	510		± 60		
VPM MPV	µm ³ ; fl	8.8	9.2		± 2.0	8.1	8.4		± 2.0	8.0	8.5		± 2.0		
LYMPHO	#	1.3	1.3		± 0.2	2.7	2.7		± 0.5	3.6	3.7		± 1.3		
	%	67.1	64.0		± 8.0	35.7	35.5		± 6.0	18.2	17.5		± 6.0		
MONO	#	0.2	0.2		± 0.2	0.5	0.5		± 0.4	1.2	0.8		± 0.8		
	%	8.4	9.0		± 7.0	7.1	6.0		± 5.0	6.0	4.0		± 4.0		
GRANULO	#	0.5	0.6		± 0.2	4.3	4.5		± 0.5	15.2	16.6		± 1.5		
	%	24.5	27.0		± 7.0	57.3	58.5		± 6.0	76.2	78.5		± 7.0		
CRP	mg/dl	0.59	N/A		± 0.30	2.11	N/A		± 0.70	4.68	N/A		± 1.00		
	mg/l	5.9	N/A		± 3.0	21.1	N/A		± 7.0	46.8	N/A		± 10.0		

Ref: TEMP-1070 Rev.38 1300027267-A