

ABX Minotrol CRP



LOT

MC 405

CONTROL



(Exp.) 2018-01-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	Minilysebio														
			CONTROL				1	CONTROL				2	CONTROL				3
			MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
			CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1		± 0.4	7.6	7.7		± 1.0	19.8	20.4		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.21	2.22		± 0.15	4.67	4.65		± 0.18	5.64	5.68		± 0.25				
HB HGB	g/dl	5.5	5.6		± 0.4	13.2	13.4		± 0.5	17.7	18.2		± 0.7				
	g/l	55	56		± 4	132	134		± 5	177	182		± 7				
HT HCT	mmol/l	3.42	3.48		± 0.25	8.22	8.32		± 0.31	10.99	11.30		± 0.44				
	%	15.5	15.5		± 2.0	37.8	37.7		± 2.5	49.6	51.1		± 3.0				
VGM MCV	l/l	0.155	0.155		± 0.020	0.378	0.377		± 0.025	0.496	0.511		± 0.030				
	µm ³ ; fl	70	70		± 4	81	81		± 4	88	90		± 4				
TGMH MCH	pg	24.9	25.2		± 2.4	28.3	28.8		± 2.4	31.4	32.0		± 2.8				
	fmol	1.55	1.57		± 0.15	1.76	1.79		± 0.15	1.95	1.99		± 0.18				
CCMH MCHC	g/dl	35.6	36.0		± 3.0	35.0	35.6		± 3.0	35.7	35.6		± 3.0				
	g/l	356	360		± 30	350	356		± 30	357	356		± 30				
	mmol/l	22.12	22.38		± 1.86	21.72	22.09		± 1.86	22.15	22.11		± 1.86				
IDR RDW	%	14.5	13.8		± 3.0	14.0	14.5		± 3.0	13.5	13.5		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	69	75		± 20	255	257		± 40	530	525		± 60				
VPM MPV	µm ³ ; fl	8.7	8.7		± 2.0	8.1	8.0		± 2.0	8.3	8.2		± 2.0				
LYMPHO	#	1.5	1.4		± 0.2	2.8	2.7		± 0.5	3.4	3.6		± 1.3				
	%	70.2	67.0		± 8.0	36.8	35.5		± 6.0	17.1	17.5		± 6.0				
MONO	#	0.1	0.1		± 0.1	0.5	0.4		± 0.4	1.1	0.8		± 0.8				
	%	6.8	6.5		± 6.5	6.7	5.5		± 5.0	5.5	4.0		± 4.0				
GRANULO	#	0.5	0.6		± 0.2	4.3	4.5		± 0.5	15.4	16.0		± 1.5				
	%	23.5	26.5		± 7.0	56.8	59.0		± 6.0	77.7	78.5		± 7.0				
CRP	mg/dl	0.46	N/A		± 0.30	1.94	N/A		± 0.70	3.27	N/A		± 1.00				
	mg/l	4.6	N/A		± 3.0	19.4	N/A		± 7.0	32.7	N/A		± 10.0				

Ref: TEMP-1070 Rev.37 1300027267-A