

ABX Minotrol CRP



LOT

MC 423

CONTROL



(Exp.) **2022-07-05**
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.2		± 0.4	7.5	7.6		± 1.0	20.0	20.8		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.22	2.19		± 0.15	4.53	4.50		± 0.18	5.67	5.64		± 0.25				
HB HGB	g/dl	5.6	5.7		± 0.4	13.0	13.3		± 0.5	18.0	18.5		± 0.7				
	g/l	56	57		± 4	130	133		± 5	180	185		± 7				
	mmol/l	3.48	3.54		± 0.25	8.09	8.26		± 0.31	11.18	11.49		± 0.44				
HT HCT	%	16.0	16.2		± 2.0	36.7	36.9		± 2.5	51.0	51.3		± 3.0				
	l/l	0.160	0.162		± 0.020	0.367	0.369		± 0.025	0.510	0.513		± 0.030				
VGM MCV	µm ³ ; fl	72	74		± 4	81	82		± 4	90	91		± 4				
TGMH MCH	pg	25.2	26.0		± 2.4	28.8	29.6		± 2.4	31.7	32.8		± 2.8				
	fmol	1.57	1.62		± 0.15	1.79	1.84		± 0.15	1.97	2.04		± 0.18				
CCMH MCHC	g/dl	35.1	35.2		± 3.0	35.5	36.0		± 3.0	35.3	36.0		± 3.0				
	g/l	351	352		± 30	355	360		± 30	353	360		± 30				
	mmol/l	21.78	21.84		± 1.86	22.06	22.38		± 1.86	21.90	22.38		± 1.86				
IDR RDW	%	13.5	14.0		± 3.0	13.0	14.0		± 3.0	12.5	13.0		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	80	80		± 20	253	251		± 40	495	502		± 60				
VPM MPV	µm ³ ; fl	8.2	9.2		± 2.0	8.1	9.3		± 2.0	7.9	9.0		± 2.0				
LYMPHO	#	1.3	1.3		± 0.2	2.6	2.5		± 0.5	3.7	3.7		± 1.1				
	%	61.9	60.0		± 8.0	34.7	33.5		± 6.0	18.7	18.0		± 5.0				
MONO	#	0.2	0.2		± 0.2	0.7	0.6		± 0.4	1.4	1.1		± 1.0				
	%	10.8	11.0		± 7.0	9.0	8.0		± 5.0	7.0	5.5		± 5.0				
GRANULO	#	0.6	0.6		± 0.2	4.2	4.4		± 0.5	14.9	15.9		± 1.5				
	%	26.5	29.0		± 7.0	56.4	58.5		± 6.0	74.7	76.5		± 7.0				
CRP	mg/dl	0.42	N/A		± 0.30	2.66	N/A		± 0.70	4.16	N/A		± 1.00				
	mg/l	4.2	N/A		± 3.0	26.6	N/A		± 7.0	41.6	N/A		± 10.0				

1300027267-A
Ref: TEMP-1070 Rev.39