

ABX Minotrol CRP



LOT

MC 413

CONTROL



(Exp.) 2020-01-05

(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	Minilysebio													TOLERANCES TOLERANCE
			CONTROL			1	CONTROL			2	CONTROL			3		
			MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l		2.0	2.0		± 0.4	7.1	7.7		± 1.0	18.7	20.4		± 2.2		
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l		2.31	2.31		± 0.15	4.50	4.54		± 0.18	5.62	5.62		± 0.25		
HB HGB	g/dl		5.8	5.9		± 0.4	13.0	13.3		± 0.5	17.5	17.7		± 0.7		
	g/l		58	59		± 4	130	133		± 5	175	177		± 7		
HT HCT	mmol/l		3.60	3.66		± 0.25	8.09	8.26		± 0.31	10.87	10.99		± 0.44		
	%		16.4	16.4		± 2.0	36.5	36.3		± 2.5	49.5	48.9		± 3.0		
VGM MCV	l/l		0.164	0.164		± 0.020	0.365	0.363		± 0.025	0.495	0.489		± 0.030		
	µm ³ ; fl		71	71		± 4	81	80		± 4	88	87		± 4		
TGMH MCH	pg		25.1	25.5		± 2.4	29.0	29.3		± 2.4	31.1	31.5		± 2.8		
	fmol		1.56	1.59		± 0.15	1.80	1.82		± 0.15	1.93	1.96		± 0.18		
CCMH MCHC	g/dl		35.3	36.0		± 3.0	35.8	36.6		± 3.0	35.4	36.2		± 3.0		
	g/l		353	360		± 30	358	366		± 30	354	362		± 30		
	mmol/l		21.94	22.34		± 1.86	22.21	22.74		± 1.86	21.97	22.48		± 1.86		
IDR RDW	%		14.0	14.8		± 3.0	13.5	14.3		± 3.0	13.0	13.5		± 3.0		
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l		77	80		± 20	250	270		± 40	490	510		± 60		
VPM MPV	µm ³ ; fl		7.9	8.2		± 2.0	7.5	7.8		± 2.0	7.3	7.5		± 2.0		
LYMPHO	#		1.4	1.4		± 0.2	2.6	2.7		± 0.5	3.7	3.6		± 1.3		
	%		71.2	68.0		± 7.0	36.8	34.8		± 6.0	19.8	17.5		± 6.0		
MONO	#		0.1	0.2		± 0.2	0.6	0.5		± 0.4	1.0	0.9		± 0.8		
	%		6.4	6.5		± 6.0	8.6	7.0		± 5.0	5.5	4.5		± 4.0		
GRANULO	#		0.5	0.5		± 0.2	3.9	4.5		± 0.5	14.1	15.9		± 1.5		
	%		23.0	25.5		± 7.0	54.9	58.2		± 6.0	75.2	78.0		± 7.0		
CRP	mg/dl		0.39	N/A		± 0.30	2.07	N/A		± 0.70	3.95	N/A		± 1.00		
	mg/l		3.9	N/A		± 3.0	20.7	N/A		± 7.0	39.5	N/A		± 10.0		

Ref: TEMP-1070 Rev.38 1300027267-A