

ABX Minotrol CRP



LOT

MC 422

CONTROL



(Exp.) **2022-04-05**
(YYYY - MM - DD)

Rev 1

		Minilysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2	2.1			± 0.4	8.6	8.7			± 1.0	20.6	21.6			± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.33	2.28			± 0.15	4.53	4.45			± 0.18	5.56	5.46			± 0.25			
HB HGB	g/dl	5.9	5.9			± 0.4	13.1	13.5			± 0.5	17.4	17.9			± 0.7			
	g/l	59	59			± 4	131	135			± 5	174	179			± 7			
	mmol/l	3.66	3.66			± 0.25	8.15	8.38			± 0.31	10.81	11.12			± 0.44			
HT HCT	%	16.8	16.9			± 2.0	37.1	37.4			± 2.5	49.5	50.2			± 3.0			
	l/l	0.168	0.169			± 0.020	0.371	0.374			± 0.025	0.495	0.502			± 0.030			
VGM MCV	µm ³ ; fl	72	74			± 4	82	84			± 4	89	92			± 4			
TGMH MCH	pg	25.3	25.9			± 2.4	29.0	30.3			± 2.4	31.3	32.8			± 2.8			
	fmol	1.57	1.61			± 0.15	1.80	1.88			± 0.15	1.94	2.04			± 0.18			
CCMH MCHC	g/dl	35.1	35.0			± 3.0	35.4	36.1			± 3.0	35.2	35.6			± 3.0			
	g/l	351	350			± 30	354	361			± 30	352	356			± 30			
	mmol/l	21.80	21.72			± 1.86	21.95	22.43			± 1.86	21.84	22.13			± 1.86			
IDR RDW	%	13.5	14.0			± 3.0	13.0	13.5			± 3.0	12.0	12.5			± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	79	72			± 20	268	269			± 40	499	495			± 60			
VPM MPV	µm ³ ; fl	8.5	9.6			± 2.0	8.0	8.9			± 2.0	8.1	9.0			± 2.0			
LYMPHO	#	1.5	1.4			± 0.2	3.0	2.9			± 0.5	3.7	3.8			± 1.1			
	%	69.2	66.5			± 8.0	35.2	33.5			± 6.0	18.2	17.5			± 5.0			
MONO	#	0.2	0.2			± 0.2	0.7	0.5			± 0.4	1.0	0.9			± 0.9			
	%	8.0	7.5			± 7.0	7.6	6.0			± 5.0	5.0	4.0			± 4.0			
GRANULO	#	0.5	0.5			± 0.2	4.9	5.3			± 0.5	15.9	17.0			± 1.5			
	%	23.0	26.0			± 7.0	57.3	60.5			± 6.0	77.2	78.5			± 7.0			
CRP	mg/dl	0.51	N/A			± 0.30	2.23	N/A			± 0.70	4.23	N/A			± 1.00			
	mg/l	5.1	N/A			± 3.0	22.3	N/A			± 7.0	42.3	N/A			± 10.0			

1300027267-A
Ref: TEMP-1070 Rev.39