

ABX Minotrol CRP

LOT

MC 402

CONTROL



(Exp.) **2017-04-05**

(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	Minilysebio														
			CONTROL				1	CONTROL				2	CONTROL				3
			MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
			CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0		± 0.4	7.6	7.5		± 1.0	19.7	20.1		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.42	2.40		± 0.15	4.72	4.68		± 0.18	5.65	5.58		± 0.25				
HB HGB	g/dl	6.0	5.9		± 0.4	13.4	13.6		± 0.5	17.8	17.8		± 0.7				
	g/l	60	59		± 4	134	136		± 5	178	178		± 7				
	mmol/l	3.72	3.66		± 0.25	8.34	8.45		± 0.31	11.05	11.05		± 0.44				
HT HCT	%	16.9	16.8		± 2.0	37.8	37.9		± 2.5	50.3	49.7		± 3.0				
	l/l	0.169	0.168		± 0.020	0.378	0.379		± 0.025	0.503	0.497		± 0.030				
VGM MCV	µm ³ ; fl	70	70		± 4	80	81		± 4	89	89		± 4				
TGMH MCH	pg	24.7	24.6		± 2.4	28.4	29.1		± 2.4	31.5	31.9		± 2.8				
	fmol	1.54	1.53		± 0.15	1.77	1.80		± 0.15	1.96	1.98		± 0.18				
CCMH MCHC	g/dl	35.3	35.1		± 3.0	35.6	35.9		± 3.0	35.4	35.8		± 3.0				
	g/l	353	351		± 30	356	359		± 30	354	358		± 30				
	mmol/l	21.94	21.81		± 1.86	22.08	22.28		± 1.86	21.98	22.26		± 1.86				
IDR RDW	%	13.8	13.7		± 3.0	13.2	14.4		± 3.0	12.8	12.9		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	70	73		± 20	255	255		± 40	510	520		± 60				
VPM MPV	µm ³ ; fl	9.0	9.0		± 2.0	8.7	8.6		± 2.0	8.6	8.2		± 2.0				
LYMPHO	#	1.3	1.3		± 0.3	2.7	2.7		± 0.6	3.5	3.6		± 1.3				
	%	66.6	66.6		± 8.0	35.7	36.0		± 6.0	17.7	17.9		± 6.0				
MONO	#	0.2	0.2		± 0.2	0.5	0.5		± 0.5	1.1	0.9		± 0.9				
	%	8.0	8.4		± 7.0	6.7	6.0		± 5.0	5.5	4.6		± 4.6				
GRANULO	#	0.5	0.5		± 0.3	4.4	4.4		± 0.6	15.2	15.6		± 1.5				
	%	25.5	25.0		± 8.0	57.8	58.0		± 6.0	77.2	77.5		± 7.0				
CRP	mg/dl	0.54	N/A		± 0.30	2.27	N/A		± 0.70	4.28	N/A		± 1.00				
	mg/l	5.4	N/A		± 3.0	22.7	N/A		± 7.0	42.8	N/A		± 10.0				

Ref: TEMP-1070 Rev.37 1300027267-A