

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 044**

**CONTROL**



(Exp.) 2014-07-05

(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				EMI CRP	CRP 200			EMI CRP	CRP 200			EMI CRP	CRP 200				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.0</b>		± 0.4	<b>7.7</b>	<b>7.4</b>		± 1.0	<b>20.5</b>	<b>19.7</b>		± 2.2				
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.43</b>	<b>2.37</b>		± 0.15	<b>4.60</b>	<b>4.60</b>		± 0.18	<b>5.62</b>	<b>5.66</b>		± 0.25				
HB HGB	g/dl	<b>6.1</b>	<b>5.8</b>		± 0.4	<b>13.4</b>	<b>13.2</b>		± 0.5	<b>17.7</b>	<b>17.9</b>		± 0.7				
	g/l	<b>61</b>	<b>58</b>		± 4	<b>134</b>	<b>132</b>		± 5	<b>177</b>	<b>179</b>		± 7				
	mmol/l	<b>3.79</b>	<b>3.60</b>		± 0.25	<b>8.32</b>	<b>8.22</b>		± 0.31	<b>10.99</b>	<b>11.12</b>		± 0.44				
HT HCT	%	<b>16.8</b>	<b>16.4</b>		± 2.0	<b>36.8</b>	<b>37.3</b>		± 2.5	<b>50.0</b>	<b>50.4</b>		± 3.0				
	l/l	<b>0.168</b>	<b>0.164</b>		± 0.020	<b>0.368</b>	<b>0.373</b>		± 0.025	<b>0.500</b>	<b>0.504</b>		± 0.030				
VGM MCV	µm <sup>3</sup> ; fl	<b>69</b>	<b>69</b>		± 4	<b>80</b>	<b>81</b>		± 4	<b>89</b>	<b>89</b>		± 4				
TGMH MCH	pg	<b>25.1</b>	<b>24.5</b>		± 2.4	<b>29.1</b>	<b>28.8</b>		± 2.4	<b>31.5</b>	<b>31.6</b>		± 2.8				
	fmol	<b>1.56</b>	<b>1.52</b>		± 0.15	<b>1.81</b>	<b>1.79</b>		± 0.15	<b>1.96</b>	<b>1.96</b>		± 0.18				
CCMH MCHC	g/dl	<b>36.4</b>	<b>35.4</b>		± 3.0	<b>36.4</b>	<b>35.5</b>		± 3.0	<b>35.4</b>	<b>35.5</b>		± 3.0				
	g/l	<b>364</b>	<b>354</b>		± 30	<b>364</b>	<b>355</b>		± 30	<b>354</b>	<b>355</b>		± 30				
	mmol/l	<b>22.59</b>	<b>22.01</b>		± 1.86	<b>22.61</b>	<b>22.05</b>		± 1.86	<b>21.98</b>	<b>22.07</b>		± 1.86				
IDR RDW	%	<b>13.2</b>	<b>13.5</b>		± 3.0	<b>12.6</b>	<b>13.4</b>		± 3.0	<b>12.1</b>	<b>13.2</b>		± 3.0				
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>82</b>	<b>79</b>		± 20	<b>250</b>	<b>255</b>		± 40	<b>500</b>	<b>500</b>		± 60				
VPM MPV	µm <sup>3</sup> ; fl	<b>8.6</b>	<b>8.8</b>		± 2.0	<b>8.4</b>	<b>8.1</b>		± 2.0	<b>8.4</b>	<b>7.8</b>		± 2.0				
LYMPHO	#	<b>1.4</b>	<b>1.4</b>		± 0.3	<b>2.7</b>	<b>2.6</b>		± 0.5	<b>3.4</b>	<b>3.2</b>		± 1.3				
	%	<b>64.5</b>	<b>68.1</b>		± 8.0	<b>34.5</b>	<b>35.7</b>		± 6.0	<b>16.5</b>	<b>16.1</b>		± 6.0				
MONO	#	<b>0.2</b>	<b>0.2</b>		± 0.2	<b>0.5</b>	<b>0.6</b>		± 0.4	<b>1.1</b>	<b>1.4</b>		± 1.0				
	%	<b>8.5</b>	<b>7.6</b>		± 7.0	<b>7.0</b>	<b>8.1</b>		± 5.0	<b>5.5</b>	<b>7.0</b>		± 5.0				
GRANULO	#	<b>0.6</b>	<b>0.5</b>		± 0.2	<b>4.5</b>	<b>4.2</b>		± 0.5	<b>16.0</b>	<b>15.2</b>		± 1.5				
	%	<b>27.0</b>	<b>24.5</b>		± 7.0	<b>58.5</b>	<b>56.4</b>		± 6.0	<b>78.0</b>	<b>77.2</b>		± 7.0				
CRP	mg/dl	<b>0.50</b>	<b>0.50</b>		± 0.30	<b>2.50</b>	<b>2.50</b>		± 0.70	<b>4.00</b>	<b>4.10</b>		± 1.00				
	mg/l	<b>5.0</b>	<b>5.0</b>		± 3.0	<b>25.0</b>	<b>25.0</b>		± 7.0	<b>40.0</b>	<b>41.0</b>		± 10.0				

9930087-B  
Ref: TEMP-1070 Rev.34