

ABX Minotrol CRP



LOT

MC 016

CONTROL



(Exp.) 2016-04-05
(YYYY - MM - DD)

Rev 3

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
					CRP 200					CRP 200						CRP 200		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0			± 0.4	7.2				± 1.0	18.9				± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.45			± 0.15	4.65				± 0.18	5.62				± 0.25		
HB	HGB	g/dl	5.8			± 0.4	13.1				± 0.5	18.0				± 0.7		
		g/l	58			± 4	131				± 5	180				± 7		
		mmol/l	3.60			± 0.25	8.15				± 0.31	11.18				± 0.44		
HT	HCT	%	16.9			± 2.0	37.2				± 2.5	50.0				± 3.0		
		l/l	0.169			± 0.020	0.372				± 0.025	0.500				± 0.030		
VGM	MCV	µm ³ ; fl	69			± 4	80				± 4	89				± 4		
TGMH	MCH	pg	23.7			± 2.4	28.2				± 2.4	32.0				± 2.8		
		fmol	1.47			± 0.15	1.75				± 0.15	1.99				± 0.18		
CCMH	MCHC	g/dl	34.3			± 3.0	35.3				± 3.0	36.0				± 3.0		
		g/l	343			± 30	353				± 30	360				± 30		
		mmol/l	21.29			± 1.86	21.92				± 1.86	22.35				± 1.86		
IDR	RDW	%	13.5			± 3.0	13.0				± 3.0	12.5				± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75			± 20	250				± 40	495				± 60		
VPM	MPV	µm ³ ; fl	9.0			± 2.0	8.0				± 2.0	7.8				± 2.0		
LYMPHO		#	1.4			± 0.4	2.6				± 0.5	3.0				± 1.2		
		%	69.7			± 8.0	35.7				± 6.0	16.1				± 6.0		
MONO		#	0.2			± 0.2	0.6				± 0.4	1.1				± 1.0		
		%	8.4			± 7.0	8.1				± 5.0	6.0				± 5.0		
GRANULO		#	0.4			± 0.4	4.1				± 0.5	14.8				± 1.4		
		%	22.1			± 8.0	56.4				± 6.0	78.2				± 7.0		
CRP		mg/dl	0.60			± 0.30	2.53				± 0.70	4.20				± 1.00		
		mg/l	6.0			± 3.0	25.3				± 7.0	42.0				± 10.0		

9930093-A
Ref: TEMP-1070 Rev.36