

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 074**

**CONTROL**



**(Exp.) 2014-10-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				EMI CRP	CRP 200			EMI CRP	CRP 200			EMI CRP	CRP 200				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.0</b>		± 0.4	<b>7.8</b>	<b>7.4</b>		± 1.0	<b>20.2</b>	<b>19.3</b>		± 2.2				
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.38</b>	<b>2.34</b>		± 0.15	<b>4.62</b>	<b>4.59</b>		± 0.18	<b>5.45</b>	<b>5.47</b>		± 0.25				
HB HGB	g/dl	<b>6.0</b>	<b>5.7</b>		± 0.4	<b>13.4</b>	<b>13.1</b>		± 0.5	<b>17.5</b>	<b>17.7</b>		± 0.7				
	g/l	<b>60</b>	<b>57</b>		± 4	<b>134</b>	<b>131</b>		± 5	<b>175</b>	<b>177</b>		± 7				
	mmol/l	<b>3.73</b>	<b>3.54</b>		± 0.25	<b>8.32</b>	<b>8.15</b>		± 0.31	<b>10.87</b>	<b>10.99</b>		± 0.44				
HT HCT	%	<b>16.4</b>	<b>16.4</b>		± 2.0	<b>37.0</b>	<b>37.2</b>		± 2.5	<b>48.5</b>	<b>49.2</b>		± 3.0				
	l/l	<b>0.164</b>	<b>0.164</b>		± 0.020	<b>0.370</b>	<b>0.372</b>		± 0.025	<b>0.485</b>	<b>0.492</b>		± 0.030				
VGM MCV	µm <sup>3</sup> ; fl	<b>69</b>	<b>70</b>		± 4	<b>80</b>	<b>81</b>		± 4	<b>89</b>	<b>90</b>		± 4				
TGMH MCH	pg	<b>25.2</b>	<b>24.4</b>		± 2.4	<b>29.0</b>	<b>28.6</b>		± 2.4	<b>32.1</b>	<b>32.4</b>		± 2.8				
	fmol	<b>1.57</b>	<b>1.51</b>		± 0.15	<b>1.80</b>	<b>1.78</b>		± 0.15	<b>1.99</b>	<b>2.01</b>		± 0.18				
CCMH MCHC	g/dl	<b>36.5</b>	<b>34.8</b>		± 3.0	<b>36.3</b>	<b>35.3</b>		± 3.0	<b>36.1</b>	<b>36.0</b>		± 3.0				
	g/l	<b>365</b>	<b>348</b>		± 30	<b>363</b>	<b>353</b>		± 30	<b>361</b>	<b>360</b>		± 30				
	mmol/l	<b>22.69</b>	<b>21.61</b>		± 1.86	<b>22.51</b>	<b>21.93</b>		± 1.86	<b>22.40</b>	<b>22.33</b>		± 1.86				
IDR RDW	%	<b>13.5</b>	<b>14.0</b>		± 3.0	<b>12.8</b>	<b>13.7</b>		± 3.0	<b>12.0</b>	<b>13.0</b>		± 3.0				
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>80</b>	<b>80</b>		± 20	<b>258</b>	<b>257</b>		± 40	<b>500</b>	<b>495</b>		± 60				
VPM MPV	µm <sup>3</sup> ; fl	<b>8.1</b>	<b>8.6</b>		± 2.0	<b>7.8</b>	<b>7.7</b>		± 2.0	<b>8.0</b>	<b>7.6</b>		± 2.0				
LYMPHO	#	<b>1.3</b>	<b>1.3</b>		± 0.2	<b>2.7</b>	<b>2.6</b>		± 0.5	<b>3.4</b>	<b>3.2</b>		± 1.1				
	%	<b>63.5</b>	<b>66.0</b>		± 8.0	<b>34.5</b>	<b>35.2</b>		± 6.0	<b>17.0</b>	<b>16.6</b>		± 5.0				
MONO	#	<b>0.2</b>	<b>0.1</b>		± 0.2	<b>0.5</b>	<b>0.6</b>		± 0.4	<b>1.0</b>	<b>1.0</b>		± 1.0				
	%	<b>8.0</b>	<b>6.8</b>		± 7.0	<b>7.0</b>	<b>7.6</b>		± 5.0	<b>5.0</b>	<b>5.0</b>		± 5.0				
GRANULO	#	<b>0.6</b>	<b>0.5</b>		± 0.2	<b>4.6</b>	<b>4.2</b>		± 0.5	<b>15.8</b>	<b>15.2</b>		± 1.5				
	%	<b>28.5</b>	<b>27.4</b>		± 7.0	<b>58.5</b>	<b>57.3</b>		± 6.0	<b>78.0</b>	<b>78.7</b>		± 7.0				
CRP	mg/dl	<b>0.55</b>	<b>0.46</b>		± 0.30	<b>2.50</b>	<b>2.32</b>		± 0.70	<b>4.10</b>	<b>3.85</b>		± 1.00				
	mg/l	<b>5.5</b>	<b>4.6</b>		± 3.0	<b>25.0</b>	<b>23.2</b>		± 7.0	<b>41.0</b>	<b>38.5</b>		± 10.0				

9930087-B  
Ref: TEMP-1070 Rev.35