

ABX Minotrol CRP



LOT

MC 045

CONTROL



(Exp.) 2015-07-05
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio													
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
					CRP 200				CRP 200				CRP 200				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l		1.9			± 0.4		7.6			± 1.0		19.8			± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l		2.52			± 0.15		4.61			± 0.18		5.77			± 0.25
HB	HGB	g/dl		5.9			± 0.4		13.1			± 0.5		18.0			± 0.7
		g/l		59			± 4		131			± 5		180			± 7
		mmol/l		3.66			± 0.25		8.15			± 0.31		11.18			± 0.44
HT	HCT	%		17.4			± 2.0		37.3			± 2.5		51.4			± 3.0
		l/l		0.174			± 0.020		0.373			± 0.025		0.514			± 0.030
VGM	MCV	µm ³ ; fl		69			± 4		81			± 4		89			± 4
TGMH	MCH	pg		23.4			± 2.4		28.5			± 2.4		31.2			± 2.8
		fmol		1.45			± 0.15		1.77			± 0.15		1.94			± 0.18
CCMH	MCHC	g/dl		33.9			± 3.0		35.2			± 3.0		35.1			± 3.0
		g/l		339			± 30		352			± 30		351			± 30
		mmol/l		21.04			± 1.86		21.84			± 1.86		21.77			± 1.86
IDR	RDW	%		14.0			± 3.0		13.5			± 3.0		13.0			± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l		70			± 20		255			± 40		505			± 60
VPM	MPV	µm ³ ; fl		9.0			± 2.0		7.7			± 2.0		7.6			± 2.0
LYMPHO		#		1.3			± 0.3		2.7			± 0.5		3.4			± 1.3
		%		68.6			± 8.0		35.7			± 6.0		17.1			± 6.0
MONO		#		0.2			± 0.2		0.6			± 0.5		1.3			± 1.0
		%		8.0			± 7.0		7.6			± 5.0		6.5			± 5.0
GRANULO		#		0.4			± 0.2		4.3			± 0.6		15.2			± 1.5
		%		23.5			± 7.0		56.8			± 6.0		76.7			± 7.0
CRP		mg/dl		0.46			± 0.30		2.47			± 0.70		4.00			± 1.00
		mg/l		4.6			± 3.0		24.7			± 7.0		40.0			± 10.0

9930093-A
Ref: TEMP-1070 Rev.36