

ABX Minotrol CRP



LOT

MC 046

CONTROL



(Exp.) 2016-07-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
					CRP 200					CRP 200						CRP 200		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0			± 0.4	7.4				± 1.0	19.7				± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.30			± 0.15	4.57				± 0.18	5.60				± 0.25		
HB	HGB	g/dl	5.8			± 0.4	13.1				± 0.5	17.8				± 0.7		
		g/l	58			± 4	131				± 5	178				± 7		
		mmol/l	3.60			± 0.25	8.15				± 0.31	11.05				± 0.44		
HT	HCT	%	16.3			± 2.0	37.0				± 2.5	49.8				± 3.0		
		l/l	0.163			± 0.020	0.370				± 0.025	0.498				± 0.030		
VGM	MCV	µm ³ ; fl	71			± 4	81				± 4	89				± 4		
TGMH	MCH	pg	25.2			± 2.4	28.7				± 2.4	31.8				± 2.8		
		fmol	1.56			± 0.15	1.78				± 0.15	1.97				± 0.18		
CCMH	MCHC	g/dl	35.5			± 3.0	35.5				± 3.0	35.7				± 3.0		
		g/l	355			± 30	355				± 30	357				± 30		
		mmol/l	22.04			± 1.86	22.03				± 1.86	22.18				± 1.86		
IDR	RDW	%	13.5			± 3.0	13.2				± 3.0	12.5				± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	65			± 20	265				± 40	510				± 60		
VPM	MPV	µm ³ ; fl	9.2			± 2.0	8.3				± 2.0	7.9				± 2.0		
LYMPHO		#	1.3			± 0.2	2.7				± 0.5	3.7				± 1.1		
		%	65.5			± 8.0	36.8				± 6.0	18.7				± 5.0		
MONO		#	0.2			± 0.2	0.5				± 0.4	1.0				± 1.0		
		%	9.6			± 7.0	7.1				± 5.0	5.0				± 5.0		
GRANULO		#	0.5			± 0.2	4.2				± 0.5	15.1				± 1.5		
		%	24.5			± 7.0	56.4				± 6.0	76.7				± 7.0		
CRP		mg/dl	0.70			± 0.30	2.34				± 0.70	4.00				± 1.00		
		mg/l	7.0			± 3.0	23.4				± 7.0	40.0				± 10.0		

9930093-A
Ref: TEMP-1070 Rev.36