

ABX Minotrol CRP



LOT

MC 408

CONTROL



(Exp.) **2018-10-05**
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1		± 0.4	7.5	7.5		± 1.0	20.0	20.4		± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.29	2.30		± 0.15	4.60	4.55		± 0.18	5.55	5.60		± 0.25			
HB	HGB	g/dl	5.7	5.8		± 0.4	13.0	13.2		± 0.5	17.8	18.2		± 0.7			
		g/l	57	58		± 4	130	132		± 5	178	182		± 7			
HT	HCT	mmol/l	3.54	3.60		± 0.25	8.09	8.20		± 0.31	11.05	11.30		± 0.44			
		%	16.3	16.3		± 2.0	37.3	36.9		± 2.5	49.4	49.8		± 3.0			
		l/l	0.163	0.163		± 0.020	0.373	0.369		± 0.025	0.494	0.498		± 0.030			
VGM	MCV	µm ³ ; fl	71	71		± 4	81	81		± 4	89	89		± 4			
TGMH	MCH	pg	24.9	25.2		± 2.4	28.3	29.0		± 2.4	32.1	32.5		± 2.8			
		fmol	1.55	1.57		± 0.15	1.76	1.80		± 0.15	1.99	2.02		± 0.18			
CCMH	MCHC	g/dl	35.1	35.5		± 3.0	35.0	35.8		± 3.0	36.0	36.5		± 3.0			
		g/l	351	355		± 30	350	358		± 30	360	365		± 30			
		mmol/l	21.77	22.06		± 1.86	21.72	22.24		± 1.86	22.38	22.68		± 1.86			
IDR	RDW	%	13.0	14.0		± 3.0	13.5	14.5		± 3.0	12.4	13.5		± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	76	80		± 20	265	265		± 40	525	525		± 60			
VPM	MPV	µm ³ ; fl	8.4	8.4		± 2.0	7.7	7.8		± 2.0	7.8	8.0		± 2.0			
LYMPHO		#	1.4	1.4		± 0.2	2.7	2.7		± 0.5	3.1	3.2		± 1.3			
		%	68.6	66.0		± 8.0	35.7	35.5		± 6.0	15.5	15.5		± 6.0			
MONO		#	0.1	0.2		± 0.2	0.5	0.4		± 0.4	1.2	0.9		± 0.9			
		%	6.4	8.0		± 7.0	6.7	5.5		± 5.0	6.0	4.5		± 4.5			
GRANULO		#	0.5	0.5		± 0.2	4.3	4.4		± 0.5	15.7	16.3		± 1.5			
		%	25.5	26.0		± 7.0	57.8	59.0		± 6.0	78.7	80.0		± 7.0			
CRP		mg/dl	0.49	N/A		± 0.30	2.28	N/A		± 0.70	4.38	N/A		± 1.00			
		mg/l	4.9	N/A		± 3.0	22.8	N/A		± 7.0	43.8	N/A		± 10.0			

Ref: TEMP-1070 Rev.37 1300027267-A