

ABX Minotrol CRP



LOT

MC 409

CONTROL



(Exp.) 2019-01-05
(YYYY - MM - DD)

Rev 1

		Minilysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.2			± 0.4	7.5	7.7			± 1.0	19.5	20.2			± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.29	2.31			± 0.15	4.63	4.65			± 0.18	5.85	5.83			± 0.25			
HB HGB	g/dl	5.8	5.9			± 0.4	13.1	13.5			± 0.5	18.1	18.5			± 0.7			
	g/l	58	59			± 4	131	135			± 5	181	185			± 7			
HT HCT	mmol/l	3.60	3.66			± 0.25	8.15	8.38			± 0.31	11.24	11.49			± 0.44			
	%	16.3	16.4			± 2.0	37.5	37.7			± 2.5	51.5	51.3			± 3.0			
VGM MCV	l/l	0.163	0.164			± 0.020	0.375	0.377			± 0.025	0.515	0.513			± 0.030			
	µm ³ ; fl	71	71			± 4	81	81			± 4	88	88			± 4			
TGMH MCH	pg	25.3	25.5			± 2.4	28.4	29.0			± 2.4	30.9	31.7			± 2.8			
	fmol	1.57	1.59			± 0.15	1.76	1.80			± 0.15	1.92	1.97			± 0.18			
CCMH MCHC	g/dl	35.6	36.0			± 3.0	35.0	35.8			± 3.0	35.2	36.1			± 3.0			
	g/l	356	360			± 30	350	358			± 30	352	361			± 30			
	mmol/l	22.13	22.34			± 1.86	21.74	22.26			± 1.86	21.83	22.39			± 1.86			
IDR RDW	%	13.5	14.0			± 3.0	13.5	14.5			± 3.0	13.0	13.5			± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	76	80			± 20	255	262			± 40	500	505			± 60			
VPM MPV	µm ³ ; fl	8.3	8.3			± 2.0	8.1	8.0			± 2.0	7.9	8.0			± 2.0			
LYMPHO	#	1.4	1.5			± 0.2	2.7	2.7			± 0.5	3.3	3.4			± 1.2			
	%	67.6	66.0			± 8.0	35.7	35.5			± 6.0	17.1	17.0			± 6.0			
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.1	1.0			± 1.0			
	%	7.2	8.0			± 7.0	7.6	6.5			± 5.0	5.5	5.0			± 5.0			
GRANULO	#	0.5	0.6			± 0.2	4.3	4.5			± 0.6	15.2	15.8			± 1.5			
	%	25.5	26.0			± 7.0	56.8	58.0			± 6.0	77.7	78.0			± 7.0			
CRP	mg/dl	0.55	N/A			± 0.30	2.70	N/A			± 0.70	4.89	N/A			± 1.00			
	mg/l	5.5	N/A			± 3.0	27.0	N/A			± 7.0	48.9	N/A			± 10.0			

Ref: TEMP-1070 Rev.37 1300027267-A