

ABX Minotrol CRP



LOT

MC 400

CONTROL



(Exp.) 2016-10-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
					CRP 200					CRP 200						CRP 200		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0			± 0.4	7.3				± 1.0	19.1				± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.35			± 0.15	4.60				± 0.18	5.76				± 0.25		
HB	HGB	g/dl	5.9			± 0.4	13.3				± 0.5	18.1				± 0.7		
		g/l	59			± 4	133				± 5	181				± 7		
		mmol/l	3.66			± 0.25	8.28				± 0.31	11.24				± 0.44		
HT	HCT	%	16.5			± 2.0	37.3				± 2.5	50.1				± 3.0		
		l/l	0.165			± 0.020	0.373				± 0.025	0.501				± 0.030		
VGM	MCV	µm ³ ; fl	70			± 4	81				± 4	87				± 4		
TGMH	MCH	pg	25.1			± 2.4	29.0				± 2.4	31.4				± 2.8		
		fmol	1.56			± 0.15	1.80				± 0.15	1.95				± 0.18		
CCMH	MCHC	g/dl	35.8			± 3.0	35.8				± 3.0	36.1				± 3.0		
		g/l	358			± 30	358				± 30	361				± 30		
		mmol/l	22.24			± 1.86	22.21				± 1.86	22.43				± 1.86		
IDR	RDW	%	13.6			± 3.0	13.2				± 3.0	12.8				± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75			± 20	254				± 40	510				± 60		
VPM	MPV	µm ³ ; fl	8.7			± 2.0	8.0				± 2.0	7.9				± 2.0		
LYMPHO		#	1.3			± 0.3	2.6				± 0.5	3.2				± 1.2		
		%	66.6			± 8.0	35.7				± 6.0	16.6				± 6.0		
MONO		#	0.2			± 0.2	0.6				± 0.4	1.1				± 1.0		
		%	8.0			± 7.0	7.6				± 5.0	6.0				± 5.0		
GRANULO		#	0.5			± 0.3	4.1				± 0.5	14.9				± 1.4		
		%	25.5			± 8.0	56.8				± 6.0	77.7				± 7.0		
CRP		mg/dl	0.49			± 0.30	2.65				± 0.70	4.75				± 1.00		
		mg/l	4.9			± 3.0	26.5				± 7.0	47.5				± 10.0		

9930093-A
Ref: TEMP-1070 Rev.36