

ABX Minotrol CRP



LOT

MC 410

CONTROL



(Exp.) 2019-04-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	Minilysebio														
			CONTROL				1	CONTROL				2	CONTROL				3
			MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
			CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.6	7.8			± 1.0	19.5	20.5			± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.37	2.33			± 0.15	4.75	4.72			± 0.18	5.73	5.70			± 0.25
HB	HGB	g/dl	5.8	5.8			± 0.4	13.2	13.5			± 0.5	18.0	18.5			± 0.7
		g/l	58	58			± 4	132	135			± 5	180	185			± 7
HT	HCT	mmol/l	3.60	3.60			± 0.25	8.22	8.38			± 0.31	11.18	11.49			± 0.44
		%	16.6	16.3			± 2.0	38.0	37.8			± 2.5	51.0	50.7			± 3.0
		l/l	0.166	0.163			± 0.020	0.380	0.378			± 0.025	0.510	0.507			± 0.030
VGM	MCV	µm ³ ; fl	70	70			± 4	80	80			± 4	89	89			± 4
TGMH	MCH	pg	24.5	24.9			± 2.4	27.9	28.6			± 2.4	31.4	32.5			± 2.8
		fmol	1.52	1.55			± 0.15	1.73	1.78			± 0.15	1.95	2.02			± 0.18
CCMH	MCHC	g/dl	34.9	35.6			± 3.0	34.8	35.8			± 3.0	35.3	36.5			± 3.0
		g/l	349	356			± 30	348	358			± 30	353	365			± 30
		mmol/l	21.69	22.08			± 1.86	21.62	22.20			± 1.86	21.92	22.65			± 1.86
IDR	RDW	%	13.0	14.0			± 3.0	13.5	14.0			± 3.0	13.0	13.5			± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	76			± 20	265	271			± 40	515	523			± 60
VPM	MPV	µm ³ ; fl	8.4	8.3			± 2.0	8.4	8.4			± 2.0	8.0	8.0			± 2.0
LYMPHO		#	1.4	1.3			± 0.2	2.7	2.7			± 0.5	3.1	3.3			± 1.2
		%	67.6	65.5			± 8.0	35.7	35.0			± 6.0	16.1	16.0			± 6.0
MONO		#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.2	1.0			± 1.0
		%	7.6	8.0			± 7.0	7.6	7.0			± 5.0	6.0	5.0			± 5.0
GRANULO		#	0.5	0.5			± 0.2	4.3	4.5			± 0.5	15.3	16.2			± 1.4
		%	25.0	26.5			± 7.0	56.8	58.0			± 6.0	78.2	79.0			± 7.0
CRP		mg/dl	0.52	N/A			± 0.30	1.95	N/A			± 0.70	3.57	N/A			± 1.00
		mg/l	5.2	N/A			± 3.0	19.5	N/A			± 7.0	35.7	N/A			± 10.0

1300027267-A
Ref: TEMP-1070 Rev.38