

ABX Minotrol CRP



LOT

MC 401

CONTROL



(Exp.) 2017-01-05

(YYYY - MM - DD)

Rev 1

		Minilysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				TOLERANCES TOLERANCE	CONTROL				TOLERANCES TOLERANCE	CONTROL				TOLERANCES TOLERANCE	
		MICROS					MICROS					MICROS					
		CRP 200					CRP 200					CRP 200					CRP 200
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1				± 0.4	7.5					± 1.0	19.6				± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.43				± 0.15	4.60					± 0.18	5.67				± 0.25
HB HGB	g/dl	6.0				± 0.4	13.3					± 0.5	18.0				± 0.7
	g/l	60				± 4	133					± 5	180				± 7
HT HCT	mmol/l	3.72				± 0.25	8.28					± 0.31	11.18				± 0.44
	%	17.0				± 2.0	37.7					± 2.5	49.9				± 3.0
VGM MCV	l/l	0.170				± 0.020	0.377					± 0.025	0.499				± 0.030
	µm ³ ; fl	70				± 4	82					± 4	88				± 4
TGMH MCH	pg	24.6				± 2.4	29.0					± 2.4	31.7				± 2.8
	fmol	1.53				± 0.15	1.80					± 0.15	1.97				± 0.18
CCMH MCHC	g/dl	35.2				± 3.0	35.3					± 3.0	36.1				± 3.0
	g/l	352				± 30	353					± 30	361				± 30
	mmol/l	21.85				± 1.86	21.94					± 1.86	22.40				± 1.86
IDR RDW	%	14.0				± 3.0	14.0				± 3.0	13.0				± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	75				± 20	246					± 40	480				± 60
VPM MPV	µm ³ ; fl	8.0				± 2.0	7.8					± 2.0	7.5				± 2.0
LYMPHO	#	1.4				± 0.3	2.6					± 0.5	3.3				± 1.3
	%	65.5				± 8.0	34.1					± 6.0	16.6				± 6.0
MONO	#	0.2				± 0.2	0.5					± 0.4	1.1				± 1.0
	%	9.6				± 7.0	7.1					± 5.0	5.5				± 5.0
GRANULO	#	0.5				± 0.3	4.4					± 0.5	15.3				± 1.5
	%	24.5				± 8.0	58.8					± 6.0	78.2				± 7.0
CRP	mg/dl	0.48				± 0.30	2.36					± 0.70	4.20				± 1.00
	mg/l	4.8				± 3.0	23.6					± 7.0	42.0				± 10.0

Ref: TEMP-1070 Rev.36 9930093-A