

ABX Minotrol CRP



LOT

MC 411

CONTROL



(Exp.) **2019-07-05**

(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	Minilysebio														
			CONTROL				1	CONTROL				2	CONTROL				3
			MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
			CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.2		± 0.4	7.4	7.6		± 1.0	19.0	20.2		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.36	2.37		± 0.15	4.55	4.58		± 0.18	5.59	5.60		± 0.25				
HB HGB	g/dl	5.7	5.7		± 0.4	13.0	13.3		± 0.5	17.7	18.2		± 0.7				
	g/l	57	57		± 4	130	133		± 5	177	182		± 7				
HT HCT	mmol/l	3.54	3.54		± 0.25	8.09	8.26		± 0.31	10.99	11.30		± 0.44				
	%	16.5	16.1		± 2.0	36.9	37.1		± 2.5	49.8	49.8		± 3.0				
VGM MCV	l/l	0.165	0.161		± 0.020	0.369	0.371		± 0.025	0.498	0.498		± 0.030				
	µm ³ ; fl	70	68		± 4	81	81		± 4	89	89		± 4				
TGMH MCH	pg	24.2	24.1		± 2.4	28.6	29.0		± 2.4	31.7	32.5		± 2.8				
	fmol	1.50	1.49		± 0.15	1.78	1.80		± 0.15	1.97	2.02		± 0.18				
CCMH MCHC	g/dl	34.5	35.4		± 3.0	35.4	35.9		± 3.0	35.6	36.5		± 3.0				
	g/l	345	354		± 30	354	359		± 30	356	365		± 30				
	mmol/l	21.43	21.96		± 1.86	21.96	22.26		± 1.86	22.09	22.68		± 1.86				
IDR RDW	%	14.5	14.4		± 3.0	13.5	14.0		± 3.0	12.8	13.5		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	75	80		± 20	265	270		± 40	505	520		± 60				
VPM MPV	µm ³ ; fl	8.6	9.0		± 2.0	8.0	8.6		± 2.0	8.0	8.4		± 2.0				
LYMPHO	#	1.4	1.4		± 0.2	2.8	2.6		± 0.5	3.4	3.2		± 1.2				
	%	67.6	65.5		± 7.0	37.3	34.5		± 6.0	17.7	16.0		± 6.0				
MONO	#	0.1	0.2		± 0.2	0.5	0.5		± 0.4	0.9	0.8		± 0.8				
	%	6.8	7.0		± 6.0	6.7	6.0		± 5.0	4.5	4.0		± 4.0				
GRANULO	#	0.5	0.6		± 0.2	4.2	4.5		± 0.5	14.9	16.2		± 1.4				
	%	26.0	27.5		± 7.0	56.4	59.5		± 6.0	78.2	80.0		± 7.0				
CRP	mg/dl	0.58	N/A		± 0.30	2.58	N/A		± 0.70	4.36	N/A		± 1.00				
	mg/l	5.8	N/A		± 3.0	25.8	N/A		± 7.0	43.6	N/A		± 10.0				

1300027267-A
Ref: TEMP-1070 Rev.38