

# ABX Minotrol CRP



**LOT**

**MC 420**

**CONTROL**



(Exp.) **2021-10-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>		± 0.4	<b>7.6</b>	<b>7.9</b>		± 1.0	<b>19.4</b>	<b>20.7</b>		± 2.2				
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.33</b>	<b>2.32</b>		± 0.15	<b>4.55</b>	<b>4.55</b>		± 0.18	<b>5.48</b>	<b>5.52</b>		± 0.25				
HB HGB	g/dl	<b>5.7</b>	<b>5.8</b>		± 0.4	<b>13.1</b>	<b>13.5</b>		± 0.5	<b>17.6</b>	<b>17.9</b>		± 0.7				
	g/l	<b>57</b>	<b>58</b>		± 4	<b>131</b>	<b>135</b>		± 5	<b>176</b>	<b>179</b>		± 7				
	mmol/l	<b>3.54</b>	<b>3.60</b>		± 0.25	<b>8.15</b>	<b>8.38</b>		± 0.31	<b>10.93</b>	<b>11.12</b>		± 0.44				
HT HCT	%	<b>16.5</b>	<b>16.5</b>		± 2.0	<b>37.8</b>	<b>37.8</b>		± 2.5	<b>49.9</b>	<b>49.7</b>		± 3.0				
	l/l	<b>0.165</b>	<b>0.165</b>		± 0.020	<b>0.378</b>	<b>0.378</b>		± 0.025	<b>0.499</b>	<b>0.497</b>		± 0.030				
VGM MCV	µm <sup>3</sup> ; fl	<b>71</b>	<b>71</b>		± 4	<b>83</b>	<b>83</b>		± 4	<b>91</b>	<b>90</b>		± 4				
TGMH MCH	pg	<b>24.5</b>	<b>25.0</b>		± 2.4	<b>28.9</b>	<b>29.7</b>		± 2.4	<b>32.1</b>	<b>32.4</b>		± 2.8				
	fmol	<b>1.52</b>	<b>1.55</b>		± 0.15	<b>1.79</b>	<b>1.84</b>		± 0.15	<b>1.99</b>	<b>2.01</b>		± 0.18				
CCMH MCHC	g/dl	<b>34.5</b>	<b>35.2</b>		± 3.0	<b>34.8</b>	<b>35.7</b>		± 3.0	<b>35.3</b>	<b>36.0</b>		± 3.0				
	g/l	<b>345</b>	<b>352</b>		± 30	<b>348</b>	<b>357</b>		± 30	<b>353</b>	<b>360</b>		± 30				
	mmol/l	<b>21.40</b>	<b>21.87</b>		± 1.86	<b>21.59</b>	<b>22.20</b>		± 1.86	<b>21.92</b>	<b>22.38</b>		± 1.86				
IDR RDW	%	<b>14.0</b>	<b>14.5</b>		± 3.0	<b>13.5</b>	<b>14.0</b>		± 3.0	<b>13.0</b>	<b>13.5</b>		± 3.0				
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>71</b>	<b>76</b>		± 20	<b>256</b>	<b>257</b>		± 40	<b>481</b>	<b>505</b>		± 60				
VPM MPV	µm <sup>3</sup> ; fl	<b>9.3</b>	<b>10.0</b>		± 2.0	<b>8.5</b>	<b>8.9</b>		± 2.0	<b>8.1</b>	<b>9.0</b>		± 2.0				
LYMPHO	#	<b>1.4</b>	<b>1.4</b>		± 0.2	<b>3.1</b>	<b>3.0</b>		± 0.5	<b>3.5</b>	<b>3.5</b>		± 1.1				
	%	<b>72.3</b>	<b>68.5</b>		± 8.0	<b>40.4</b>	<b>37.5</b>		± 6.0	<b>18.2</b>	<b>17.0</b>		± 5.0				
MONO	#	<b>0.1</b>	<b>0.2</b>		± 0.2	<b>0.5</b>	<b>0.4</b>		± 0.4	<b>1.2</b>	<b>0.9</b>		± 0.9				
	%	<b>6.0</b>	<b>5.5</b>		± 5.5	<b>6.2</b>	<b>5.0</b>		± 5.0	<b>6.0</b>	<b>4.5</b>		± 4.5				
GRANULO	#	<b>0.5</b>	<b>0.5</b>		± 0.2	<b>4.1</b>	<b>4.5</b>		± 0.5	<b>14.8</b>	<b>16.2</b>		± 1.5				
	%	<b>22.5</b>	<b>26.0</b>		± 7.0	<b>53.9</b>	<b>57.5</b>		± 6.0	<b>76.2</b>	<b>78.5</b>		± 7.0				
CRP	mg/dl	<b>0.62</b>	N/A		± 0.30	<b>2.38</b>	N/A		± 0.70	<b>3.95</b>	N/A		± 1.00				
	mg/l	<b>6.2</b>	N/A		± 3.0	<b>23.8</b>	N/A		± 7.0	<b>39.5</b>	N/A		± 10.0				

1300027267-A  
Ref: TEMP-1070 Rev.39