

ABX Minotrol CRP



LOT MC 042

CONTROL



(Exp.) 2012-07-05
(YYYY - MM - DD)

Rev 3

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																							
				CONTROL				1				CONTROL				2				CONTROL				3			
				MICROS				TOLERANCES TOLERANCE				MICROS				TOLERANCES TOLERANCE				MICROS				TOLERANCES TOLERANCE			
				EMICRP								EMICRP								EMICRP							
GB	WBC	$10^3/mm^3$; $10^9/l$	2.0					± 0.4	7.2					± 1.0	19.6					± 2.2							
GR	RBC	$10^9/mm^3$; $10^{12}/l$	2.29					± 0.15	4.51					± 0.18	5.60					± 0.25							
HB	HGB	g/dl	6.0					± 0.4	13.3					± 0.5	17.9					± 0.7							
		g/l	60					± 4	133					± 5	179					± 7							
		mmol/l	3.73					± 0.25	8.26					± 0.31	11.12					± 0.44							
HT	HCT	%	16.0					± 2.0	37.0					± 2.5	51.0					± 3.0							
		l/l	0.160					± 0.020	0.370					± 0.025	0.510					± 0.030							
VGM	MCV	μm^3 ; fl	70					± 4	82					± 4	91					± 4							
TGMH	MCH	pg	26.2					± 2.4	29.5					± 2.4	32.0					± 2.8							
		fmol	1.63					± 0.15	1.83					± 0.15	1.98					± 0.18							
CCMH	MCHC	g/dl	37.4					± 3.0	36.0					± 3.0	35.1					± 3.0							
		g/l	374					± 30	360					± 30	351					± 30							
		mmol/l	23.24					± 1.86	22.33					± 1.86	21.81					± 1.86							
IDR	RDW	%	13.0					± 3.0	12.2					± 3.0	12.1					± 3.0							
PLAQ.	PLTS	$10^3/mm^3$; $10^9/l$	68					± 20	229					± 40	442					± 60							
VPM	MPV	μm^3 ; fl	7.9					± 2.0	7.9					± 2.0	7.5					± 2.0							
LYMPHO		#	1.2					± 0.2	2.3					± 0.5	2.9					± 1.1							
		%	61.2					± 7.0	32.2					± 6.0	14.7					± 5.0							
MONO		#	0.2					± 0.2	0.5					± 0.4	1.2					± 1.0							
		%	10.1					± 7.0	7.3					± 5.0	6.2					± 5.0							
GRANULO		#	0.6					± 0.2	4.4					± 0.5	15.5					± 1.5							
		%	28.7					± 7.0	60.5					± 6.0	79.1					± 7.0							
CRP		mg/dl	0.53					± 0.30	2.08					± 0.70	4.00					± 1.00							
		mg/l	5.3					± 3.0	20.8					± 7.0	40.0					± 10.0							

9930064-A

Ref: TEMP-1070 Rev.15