

ABX Minotrol CRP



LOT

MC 404

Rev 2

CONTROL



(Exp.) **2017-10-05**

(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | | | | | | |
|--------------------------|---|--------------|--|--------------|--|-------------------------|--------------|--|--------------|--|-------------------------|--------------|--|--------------|--|-------------------------|
| | | CONTROL | | | | 1 | CONTROL | | | | 2 | CONTROL | | | | 3 |
| | | MICROS | | PENTRA | | TOLERANCES TOLERANCE | MICROS | | PENTRA | | TOLERANCES TOLERANCE | MICROS | | PENTRA | | TOLERANCES TOLERANCE |
| | | EMI CRP | | MS CRP | | | EMI CRP | | MS CRP | | | EMI CRP | | MS CRP | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.1 | | 2.1 | | ± 0.4 | 7.7 | | 8.0 | | ± 1.0 | 20.4 | | 21.5 | | ± 2.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.43 | | 2.44 | | ± 0.15 | 4.68 | | 4.80 | | ± 0.18 | 5.63 | | 5.75 | | ± 0.25 |
| HB HGB | g/dl | 6.3 | | 6.1 | | ± 0.4 | 13.3 | | 13.5 | | ± 0.5 | 17.7 | | 18.3 | | ± 0.7 |
| | g/l | 63 | | 61 | | ± 4 | 133 | | 135 | | ± 5 | 177 | | 183 | | ± 7 |
| HT HCT | mmol/l | 3.91 | | 3.79 | | ± 0.25 | 8.26 | | 8.38 | | ± 0.31 | 10.99 | | 11.36 | | ± 0.44 |
| | % | 17.0 | | 16.8 | | ± 2.0 | 37.4 | | 37.7 | | ± 2.5 | 49.5 | | 49.5 | | ± 3.0 |
| VGM MCV | l/l | 0.170 | | 0.168 | | ± 0.020 | 0.374 | | 0.377 | | ± 0.025 | 0.495 | | 0.495 | | ± 0.030 |
| | μm ³ ; fl | 70 | | 69.0 | | ± 4 | 80 | | 78.5 | | ± 4 | 88 | | 86.0 | | ± 4 |
| TGMH MCH | pg | 25.9 | | 25.0 | | ± 2.4 | 28.4 | | 28.1 | | ± 2.4 | 31.4 | | 31.8 | | ± 2.8 |
| | fmol | 1.61 | | 1.55 | | ± 0.15 | 1.76 | | 1.75 | | ± 0.15 | 1.95 | | 1.98 | | ± 0.18 |
| CCMH MCHC | g/dl | 37.0 | | 36.2 | | ± 3.0 | 35.5 | | 35.8 | | ± 3.0 | 35.7 | | 37.0 | | ± 3.0 |
| | g/l | 370 | | 362 | | ± 30 | 355 | | 358 | | ± 30 | 357 | | 370 | | ± 30 |
| IDR RDW | mmol/l | 23.00 | | 22.50 | | ± 1.86 | 22.06 | | 22.25 | | ± 1.86 | 22.19 | | 22.98 | | ± 1.86 |
| | % | 13.0 | | 14.5 | | ± 3.0 | 12.5 | | 13.5 | | ± 3.0 | 12.0 | | 12.0 | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 80 | | 69 | | ± 20 | 252 | | 250 | | ± 40 | 475 | | 505 | | ± 60 |
| VPM MPV | μm ³ ; fl | 7.8 | | 9.2 | | ± 2.0 | 8.0 | | 9.2 | | ± 2.0 | 7.5 | | 8.0 | | ± 2.0 |
| LYMPHO | # | 1.3 | | N/A | | ± 0.2 | 2.7 | | N/A | | ± 0.5 | 3.7 | | N/A | | ± 1.3 |
| | % | 64.0 | | N/A | | ± 7.0 | 35.0 | | N/A | | ± 6.0 | 18.0 | | N/A | | ± 6.0 |
| MONO | # | 0.2 | | N/A | | ± 0.2 | 0.7 | | N/A | | ± 0.4 | 1.4 | | N/A | | ± 0.9 |
| | % | 7.5 | | N/A | | ± 6.0 | 9.0 | | N/A | | ± 5.0 | 7.0 | | N/A | | ± 4.5 |
| GRANULO | # | 0.6 | | N/A | | ± 0.2 | 4.3 | | N/A | | ± 0.5 | 15.3 | | N/A | | ± 1.5 |
| | % | 28.5 | | N/A | | ± 8.0 | 56.0 | | N/A | | ± 6.0 | 75.0 | | N/A | | ± 7.0 |
| CRP | mg/dl | 0.68 | | 0.67 | | ± 0.40 | 2.59 | | 2.35 | | ± 0.80 | 4.99 | | 4.65 | | ± 1.50 |
| | mg/l | 6.8 | | 6.7 | | ± 4.0 | 25.9 | | 23.5 | | ± 8.0 | 49.9 | | 46.5 | | ± 15.0 |

Ref: TEMP-1070 Rev.37 BACKVERSO 9930089-A